

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS032023\
 Data File : PS022430.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 20 Mar 2023 13:37
 Operator : AR\AJ
 Sample : HSTDICC500
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 20 15:28:20 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS032023.M
 Quant Title : 8080.M
 QLast Update : Mon Mar 20 15:26:25 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	6.299	6.722	1227.7E6	1030.8E6	511.708	501.355
Target Compounds							
1) T	Dalapon	2.406	2.479	1403.3E6	1462.4E6	450.596	446.851
2) T	3,5-DICHL...	5.681	5.922	1749.8E6	1502.5E6	474.404	470.057
3) T	4-Nitroph...	6.147	6.352	616.4E6	629.6E6	444.129	450.257
5) T	DICAMBA	6.434	6.878	4899.2E6	4697.7E6	476.851	474.588
6) T	MCPD	6.583	6.973	349.6E6	355.3E6	45.223	44.948
7) T	MCPA	6.704	7.183	477.4E6	479.7E6	45.424	45.442
8) T	DICHLORPROP	7.018	7.499	1250.2E6	1168.7E6	479.032	474.830
9) T	2,4-D	7.222	7.791	1571.7E6	1388.1E6	480.783	471.222
10) T	Pentachlo...	7.438	8.206	17950.2E6	16568.2E6	505.484	504.014
11) T	2,4,5-TP ...	7.965	8.571	7290.5E6	6626.7E6	491.110	488.431
12) T	2,4,5-T	8.226	8.946	7995.4E6	6082.8E6	492.124	480.553
13) T	2,4-DB	8.732	9.452	1453.4E6	768.4E6	482.140	447.830
14) T	DINOSEB	9.770	9.772	6134.1E6	4509.4E6	480.208	469.537
15) T	Picloram	9.614	10.744	12012.0E6	9738.0E6	484.330	480.424
16) T	DCPA	10.038	10.696	10245.6E6	7827.5E6	504.349	497.927

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS032023\
Data File : PS022430.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 20 Mar 2023 13:37
Operator : AR\AJ
Sample : HSTDICC500
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 20 15:28:20 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS032023.M
Quant Title : 8080.M
QLast Update : Mon Mar 20 15:26:25 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

