

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS032025\
 Data File : PS029472.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 20 Mar 2025 12:31
 Operator : AR\AJ
 Sample : HSTDICC500
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDICC500

Manual Integrations
 APPROVED

Reviewed By : Abdul Mirza 03/21/2025
 Supervised By : mohammad ahmed 03/24/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 21 04:39:13 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS032025.M
 Quant Title : 8080.M
 QLast Update : Fri Mar 21 04:38:06 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.159	7.600	2197.2E6	444.4E6	525.302	510.680
Target Compounds							
1) T	Dalapon	2.589	2.628	2312.4E6	891.5E6	389.097	469.442
2) T	3,5-DICHL...	6.340	6.578	2865.0E6	639.8E6	488.178	478.594
3) T	4-Nitroph...	6.963	7.136	817.1E6	334.5E6	436.044	462.753
5) T	DICAMBA	7.339	7.791	8968.3E6	2392.2E6	493.217	468.704
6) T	MCPD	7.517	7.894	596.8E6	105.9E6	45.679	46.069
7) T	MCPA	7.664	8.131	790.0E6	137.2E6	46.205	45.819
8) T	DICHLORPROP	8.040	8.497	2333.7E6	583.9E6	499.298	481.991
9) T	2,4-D	8.274	8.821	2231.6E6	579.6E6	485.248	482.843
10) T	Pentachlo...	8.560	9.331	33885.2E6	11513.7E6	548.341	489.690
11) T	2,4,5-TP ...	9.136	9.709	12731.5E6	4344.4E6	505.574	480.997
12) T	2,4,5-T	9.429	10.122	12682.2E6	3707.7E6	500.031	474.400
13) T	2,4-DB	10.000	10.683	1716.0E6	354.0E6	463.679	472.997
14) T	DINOSEB	11.187	11.055	8787.0E6	2723.9E6	500.354	475.719
15) T	Picloram	11.008	12.123	14647.9E6	5207.2E6	482.481	460.567
16) T	DCPA	11.484	12.082	15775.7E6	5844.0E6	514.209	484.356

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS032025\
Data File : PS029472.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 20 Mar 2025 12:31
Operator : AR\AJ
Sample : HSTDICC500
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDICC500

Manual Integrations
APPROVED

Reviewed By : Abdul Mirza 03/21/2025
Supervised By : mohammad ahmed 03/24/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 21 04:39:13 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS032025.M
Quant Title : 8080.M
QLast Update : Fri Mar 21 04:38:06 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

