

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS032025\
 Data File : PS029475.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 20 Mar 2025 13:43
 Operator : AR\AJ
 Sample : HSTDICC1500
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDICC1500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 21 04:41:16 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS032025.M
 Quant Title : 8080.M
 QLast Update : Fri Mar 21 04:38:06 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.157	7.599	5883.0E6	1293.6E6	1406.469	1486.573
Target Compounds							
1) T	Dalapon	2.589	2.630	7722.4E6	2851.0E6	1299.424	1501.373
2) T	3,5-DICHL...	6.338	6.576	7754.6E6	1823.9E6	1321.335	1364.350
3) T	4-Nitroph...	6.955	7.132	2733.3E6	975.5E6	1458.694	1349.343
5) T	DICAMBA	7.339	7.791	23873.6E6	7329.5E6	1312.948	1436.095
6) T	MCP	7.525	7.902	1946.1E6	333.9E6	148.965	145.193
7) T	MCPA	7.674	8.140	2470.4E6	427.2E6	144.480	142.687
8) T	DICHLORPROP	8.039	8.496	6137.8E6	1687.4E6	1313.211	1392.899
9) T	2,4-D	8.269	8.819	6067.5E6	1676.6E6	1319.319	1396.842
10) T	Pentachlo...	8.571	9.332	55916.6E6	31554.1E6	904.862	1342.030 #
11) T	2,4,5-TP ...	9.135	9.709	32545.0E6	12623.7E6	1292.376	1397.645
12) T	2,4,5-T	9.426	10.120	33181.5E6	11142.9E6	1308.277	1425.730
13) T	2,4-DB	9.995	10.681	5275.3E6	1100.1E6	1425.423	1469.909
14) T	DINOSEB	11.188	11.055	23119.9E6	8212.0E6	1316.502	1434.211
15) T	Picloram	11.002	12.121	42216.7E6	16746.9E6	1390.562	1481.224
16) T	DCPA	11.485	12.083	39328.4E6	16944.5E6	1281.905	1404.364

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS032025\
Data File : PS029475.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 20 Mar 2025 13:43
Operator : AR\AJ
Sample : HSTDICC1500
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDICC1500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 21 04:41:16 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS032025.M
Quant Title : 8080.M
QLast Update : Fri Mar 21 04:38:06 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

