

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD\_S\Data\PS032221\  
 Data File : PS014315.D  
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch  
 Acq On : 22 Mar 2021 19:45  
 Operator : DD\AJ  
 Sample : HSTDCCC750  
 Misc :  
 ALS Vial : 3 Sample Multiplier: 1

Instrument :  
 ECD\_S  
 ClientSampleId :  
 HSTDCCC750

Integration File signal 1: autoint1.e  
 Integration File signal 2: autoint2.e  
 Quant Time: Mar 23 04:43:02 2021  
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD\_S\Method\PS031721.M  
 Quant Title : 8080.M  
 QLast Update : Thu Mar 18 02:23:19 2021  
 Response via : Initial Calibration  
 Integrator: ChemStation

Volume Inj. : 2 µl  
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2  
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

|                             | Compound     | RT#1   | RT#2   | Resp#1    | Resp#2   | ng/ml   | ng/ml   |
|-----------------------------|--------------|--------|--------|-----------|----------|---------|---------|
| -----                       |              |        |        |           |          |         |         |
| System Monitoring Compounds |              |        |        |           |          |         |         |
| 4) S                        | 2,4-DCAA     | 7.164  | 6.887  | 1083.3E6  | 360.1E6  | 758.163 | 763.328 |
| Target Compounds            |              |        |        |           |          |         |         |
| 1) T                        | Dalapon      | 3.008  | 2.601  | 1080.4E6  | 413.7E6  | 694.115 | 698.668 |
| 2) T                        | 3,5-DICHL... | 6.389  | 6.048  | 1364.8E6  | 521.8E6  | 706.956 | 700.786 |
| 3) T                        | 4-Nitroph... | 6.997  | 6.501  | 549.1E6   | 208.6E6  | 735.317 | 686.836 |
| 5) T                        | DICAMBA      | 7.352  | 7.053  | 3923.7E6  | 1461.4E6 | 729.608 | 729.478 |
| 6) T                        | MCPD         | 7.520  | 7.149  | 304.0E6   | 120.7E6  | 73.108  | 72.506  |
| 7) T                        | MCPA         | 7.668  | 7.365  | 428.3E6   | 172.2E6  | 74.442  | 70.687  |
| 8) T                        | DICHLORPROP  | 8.038  | 7.678  | 951.4E6   | 366.4E6  | 718.452 | 719.042 |
| 9) T                        | 2,4-D        | 8.266  | 7.973  | 1073.1E6  | 406.5E6  | 728.045 | 680.741 |
| 10) T                       | Pentachlo... | 8.573  | 8.399  | 14298.0E6 | 4889.5E6 | 718.324 | 715.933 |
| 11) T                       | 2,4,5-TP ... | 9.118  | 8.759  | 5563.0E6  | 2000.4E6 | 721.960 | 735.253 |
| 12) T                       | 2,4,5-T      | 9.402  | 9.139  | 5625.9E6  | 1965.4E6 | 723.043 | 737.204 |
| 13) T                       | 2,4-DB       | 9.950  | 9.644  | 999.5E6   | 257.7E6  | 646.322 | 726.729 |
| 14) T                       | DINOSEB      | 11.104 | 9.974  | 4251.3E6  | 1393.3E6 | 743.793 | 771.044 |
| 15) T                       | Picloram     | 10.927 | 10.957 | 7335.7E6  | 2620.8E6 | 702.498 | 690.837 |
| 16) T                       | DCPA         | 11.369 | 10.900 | 8499.6E6  | 3017.6E6 | 721.636 | 735.669 |
| -----                       |              |        |        |           |          |         |         |

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD\_S\Data\PS032221\  
Data File : PS014315.D  
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch  
Acq On : 22 Mar 2021 19:45  
Operator : DD\AJ  
Sample : HSTDCCC750  
Misc :  
ALS Vial : 3 Sample Multiplier: 1

Instrument :  
ECD\_S  
ClientSampleId :  
HSTDCCC750

Integration File signal 1: autoint1.e  
Integration File signal 2: autoint2.e  
Quant Time: Mar 23 04:43:02 2021  
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD\_S\Method\PS031721.M  
Quant Title : 8080.M  
QLast Update : Thu Mar 18 02:23:19 2021  
Response via : Initial Calibration  
Integrator: ChemStation

Volume Inj. : 2 µl  
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2  
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

