

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS032222\
 Data File : PS018682.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Mar 2022 18:36
 Operator : AJ\MA
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 23 02:13:17 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS031422.M
 Quant Title : 8080.M
 QLast Update : Mon Mar 14 11:41:05 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.228	6.809	2063.9E6	521.3E6	725.201	740.794
Target Compounds							
1) T	Dalapon	3.029	2.534	2551.9E6	550.2E6	656.280	665.778
2) T	3,5-DICHL...	6.438	5.982	2617.4E6	718.1E6	679.204	686.492
3) T	4-Nitroph...	7.056	6.420	1069.9E6	266.5E6	620.915	655.963
5) T	DICAMBA	7.422	6.977	7525.5E6	2089.1E6	689.991	701.384
6) T	MCPD	7.594	7.075	587.3E6	165.1E6	69.193	67.427
7) T	MCPA	7.744	7.289	828.3E6	240.3E6	68.486	68.823
8) T	DICHLORPROP	8.117	7.600	1910.2E6	573.8E6	693.507	724.850
9) T	2,4-D	8.344	7.890	2330.4E6	659.7E6	695.515	704.849
10) T	Pentachlo...	8.657	8.321	26059.9E6	6994.8E6	719.835	712.282
11) T	2,4,5-TP ...	9.209	8.683	10906.1E6	3140.3E6	717.117	719.913
12) T	2,4,5-T	9.492	9.059	10903.4E6	3014.3E6	729.876	727.709
13) T	2,4-DB	10.045	9.566	2076.7E6	413.0E6	759.207	727.459
14) T	DINOSEB	11.211	9.894	8030.5E6	2188.5E6	750.999	714.307
15) T	Picloram	11.025	10.873	15581.4E6	3884.8E6	758.007	730.757
16) T	DCPA	11.478	10.828	13561.2E6	3808.2E6	757.998	724.096

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS032222\
Data File : PS018682.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Mar 2022 18:36
Operator : AJ\MA
Sample : HSTDCCC750
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 23 02:13:17 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS031422.M
Quant Title : 8080.M
QLast Update : Mon Mar 14 11:41:05 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

