

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS032222\
 Data File : PS018704.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 23 Mar 2022 05:19
 Operator : AJ\MA
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 23 06:55:46 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS031422.M
 Quant Title : 8080.M
 QLast Update : Mon Mar 14 11:41:05 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.229	6.809	2079.7E6	529.5E6	730.772	752.461
Target Compounds							
1) T	Dalapon	3.028	2.536	2532.3E6	564.4E6	651.228	682.954
2) T	3,5-DICHL...	6.437	5.982	2643.3E6	738.0E6	685.942	705.430
3) T	4-Nitroph...	7.056	6.420	1041.1E6	287.4E6	604.169	707.466
5) T	DICAMBA	7.421	6.976	7613.7E6	2146.1E6	698.075	720.533
6) T	MCPD	7.594	7.075	598.2E6	172.3E6	70.477	70.382
7) T	MCPA	7.743	7.289	841.0E6	247.8E6	69.539	70.947
8) T	DICHLORPROP	8.117	7.599	1927.0E6	577.2E6	699.626	729.137
9) T	2,4-D	8.344	7.889	2357.7E6	673.4E6	703.664	719.419
10) T	Pentachlo...	8.656	8.321	26498.8E6	7119.5E6	731.958	724.978
11) T	2,4,5-TP ...	9.209	8.681	11041.0E6	3183.5E6	725.987	729.820
12) T	2,4,5-T	9.492	9.059	11057.4E6	3016.2E6	740.186	728.162
13) T	2,4-DB	10.045	9.565	2087.2E6	399.1E6	763.037	702.902
14) T	DINOSEB	11.211	9.895	7982.1E6	2086.7E6	746.474	681.077
15) T	Picloram	11.025	10.872	15437.8E6	3728.4E6	751.023	701.344
16) T	DCPA	11.477	10.827	13716.6E6	3680.9E6	766.685	699.881

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS032222\
Data File : PS018704.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Mar 2022 05:19
Operator : AJ\MA
Sample : HSTDCCC750
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 23 06:55:46 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS031422.M
Quant Title : 8080.M
QLast Update : Mon Mar 14 11:41:05 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

