

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS032320\
 Data File : PS009403.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 23 Mar 2020 17:43
 Operator : DD\AJ
 Sample : L1749-01MS
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 BU-02-031920MS

Manual Integrations
 APPROVED

Sohil
 3/24/2020 2:52:26 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 23 18:01:27 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS031120.M
 Quant Title : 8080.M
 QLast Update : Thu Mar 12 07:13:24 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.242	6.932	194.3E6	179.8E6	351.316	327.410
Target Compounds							
1) T	Dalapon	3.047	2.631	220.9E6	222.8E6	283.232	283.031m
2) T	3,5-DICHL...	6.451	6.078	245.7E6	233.8E6	325.506	287.437
3) T	4-Nitroph...	7.071	6.534	94389289	105.7E6	395.409	332.033
5) T	DICAMBA	7.434	7.102	722.3E6	684.9E6	339.380	298.362
6) T	MCPD	7.601	7.196	52281815	48238449	21.247	26.522
7) T	MCPA	7.750	7.407	80709701	130.4E6	26.500	48.310m#
8) T	DICHLORPROP	8.128	7.730	166.2E6	185.2E6	322.293	333.862
9) T	2,4-D	8.357	8.024	213.2E6	241.5E6	404.461	367.454
10) T	Pentachlo...	8.670	8.456	2023.2E6	1964.8E6	287.744	258.862
11) T	2,4,5-TP ...	9.219	8.817	811.6E6	817.4E6	296.083	274.198
12) T	2,4,5-T	9.504	9.196	772.5E6	865.5E6	303.689	293.180
13) T	2,4-DB	10.058	9.704	85584254	108.5E6	247.386	285.701
14) T	DINOSEB	11.222	10.039	204.9E6	192.6E6	100.213	87.687
15) T	Picloram	11.035	11.018	1078.4E6	1348.4E6	355.333	331.117
16) T	DCPA	11.487	10.969	863.8E6	1037.4E6	249.204	267.845

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS032320\
Data File : PS009403.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Mar 2020 17:43
Operator : DD\AJ
Sample : L1749-01MS
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampled :
BU-02-031920MS

Manual Integrations
APPROVED

Sohil
3/24/2020 2:52:26 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 23 18:01:27 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS031120.M
Quant Title : 8080.M
QLast Update : Thu Mar 12 07:13:24 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

