

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS032320\
 Data File : PS009404.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 23 Mar 2020 18:06
 Operator : DD\AJ
 Sample : L1749-01MSD
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 BU-02-031920MSD

Manual Integrations
 APPROVED

Sohil
 3/24/2020 2:52:27 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 23 18:23:52 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS031120.M
 Quant Title : 8080.M
 QLast Update : Thu Mar 12 07:13:24 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.242	6.932	195.3E6	180.0E6	353.024	327.788
Target Compounds							
1) T	Dalapon	3.047	2.631	226.8E6	209.6E6	290.865m	266.202m
2) T	3,5-DICHL...	6.451	6.078	246.8E6	233.0E6	326.952	286.452
3) T	4-Nitroph...	7.071	6.534	95038941	106.0E6	398.131	333.125
5) T	DICAMBA	7.434	7.102	726.3E6	686.0E6	341.276	298.852
6) T	MCPD	7.600	7.195	54079935	48201716	21.978m	26.502
7) T	MCPA	7.750	7.407	81344467	134.4E6	26.708	49.815m#
8) T	DICHLORPROP	8.128	7.730	167.1E6	183.6E6	324.091	330.868
9) T	2,4-D	8.356	8.023	214.8E6	241.3E6	407.461	367.135
10) T	Pentachlo...	8.670	8.456	2015.4E6	1973.5E6	286.641	260.014
11) T	2,4,5-TP ...	9.220	8.817	815.4E6	815.4E6	297.465	273.522
12) T	2,4,5-T	9.503	9.196	778.4E6	867.2E6	305.985	293.766
13) T	2,4-DB	10.057	9.704	86728004	110.3E6	250.692	290.594
14) T	DINOSEB	11.221	10.039	206.6E6	195.7E6	101.068	89.102
15) T	Picloram	11.035	11.018	1097.3E6	1362.3E6	361.545	334.546
16) T	DCPA	11.487	10.968	869.5E6	1037.8E6	250.838	267.948

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS032320\
Data File : PS009404.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Mar 2020 18:06
Operator : DD\AJ
Sample : L1749-01MSD
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampled :
BU-02-031920MSD

Manual Integrations
APPROVED

Sohil
3/24/2020 2:52:27 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 23 18:23:52 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS031120.M
Quant Title : 8080.M
QLast Update : Thu Mar 12 07:13:24 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

