

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS032323\
 Data File : PS022491.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 23 Mar 2023 20:24
 Operator : AR\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 24 01:57:26 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS032023.M
 Quant Title : 8080.M
 QLast Update : Mon Mar 20 15:46:44 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	6.295	6.718	1832.5E6	1677.7E6	749.670	814.543
Target Compounds							
1) T	Dalapon	2.410	2.482	2073.7E6	2174.8E6	664.947	666.401
2) T	3,5-DICHL...	5.677	5.919	2605.4E6	2330.6E6	699.751	730.849
3) T	4-Nitroph...	6.137	6.343	1078.4E6	1058.9E6	787.051	754.204
5) T	DICAMBA	6.432	6.876	7426.1E6	7303.4E6	722.368	740.332
6) T	MCPD	6.585	6.976	559.8E6	582.6E6	73.630	74.844
7) T	MCPA	6.707	7.187	761.2E6	774.0E6	72.931	73.667
8) T	DICHLORPROP	7.014	7.496	1886.4E6	1848.6E6	711.756	742.395
9) T	2,4-D	7.214	7.785	2266.3E6	2186.8E6	691.150	739.153
10) T	Pentachlo...	7.437	8.205	25843.0E6	24312.8E6	734.485	740.695
11) T	2,4,5-TP ...	7.962	8.567	10861.8E6	10138.1E6	726.660	744.253
12) T	2,4,5-T	8.220	8.941	11417.2E6	9734.1E6	695.275	773.790
13) T	2,4-DB	8.724	9.446	2268.8E6	1322.1E6	752.430	787.310
14) T	DINOSEB	9.770	9.770	8994.5E6	7025.6E6	694.031	732.507
15) T	Picloram	9.607	10.738	17244.2E6	14175.2E6	694.734	706.392
16) T	DCPA	10.037	10.695	15024.8E6	11875.0E6	728.430	748.631

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS032323\
Data File : PS022491.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Mar 2023 20:24
Operator : AR\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 24 01:57:26 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS032023.M
Quant Title : 8080.M
QLast Update : Mon Mar 20 15:46:44 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

