

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS032423\
 Data File : PS022503.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 24 Mar 2023 11:09
 Operator : AR\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Manual Integrations
 APPROVED

Reviewed By :Abdul Mirza 03/27/2023
 Supervised By :Ankita Jodhani 03/27/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 24 12:30:05 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS032023.M
 Quant Title : 8080.M
 QLast Update : Mon Mar 20 15:46:44 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	6.299	6.716	1816.4E6	1676.5E6	743.102	813.948
Target Compounds							
1) T	Dalapon	2.405	2.474	2158.3E6	2303.2E6	692.075	705.745
2) T	3,5-DICHL...	5.680	5.917	2595.8E6	2360.9E6	697.166	740.333m
3) T	4-Nitroph...	6.141	6.342	1069.4E6	1052.7E6	780.490	749.774
5) T	DICAMBA	6.436	6.875	7367.0E6	7335.3E6	716.612	743.575
6) T	MCPD	6.589	6.975	551.5E6	588.5E6	72.541	75.605
7) T	MCPA	6.712	7.185	752.6E6	772.7E6	72.107	73.546
8) T	DICHLORPROP	7.019	7.495	1876.6E6	1820.5E6	708.087	731.134
9) T	2,4-D	7.221	7.784	2251.0E6	2185.8E6	686.485	738.815
10) T	Pentachlo...	7.442	8.204	25941.0E6	24242.8E6	737.270	738.562
11) T	2,4,5-TP ...	7.969	8.568	10828.5E6	10160.8E6	724.434	745.921
12) T	2,4,5-T	8.227	8.942	11497.9E6	9738.2E6	700.188	774.112
13) T	2,4-DB	8.732	9.446	2305.9E6	1299.5E6	764.731	773.808
14) T	DINOSEB	9.776	9.770	9020.5E6	7064.4E6	696.036	736.553
15) T	Picloram	9.617	10.740	17663.0E6	14580.1E6	711.605	726.572
16) T	DCPA	10.044	10.695	14897.2E6	11876.4E6	722.243	748.722

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS032423\
Data File : PS022503.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 24 Mar 2023 11:09
Operator : AR\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Manual Integrations
APPROVED

Reviewed By : Abdul Mirza 03/27/2023
Supervised By : Ankita Jodhani 03/27/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 24 12:30:05 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS032023.M
Quant Title : 8080.M
QLast Update : Mon Mar 20 15:46:44 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

