

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS032519\
 Data File : PS003722.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 25 Mar 2019 17:35
 Operator : EI\SJ
 Sample : K2002-01MS
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 S-1(2-10)-COMPMS

Manual Integrations
 APPROVED

Sohil
 3/26/2019 1:43:44 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 26 05:11:09 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS022719.M
 Quant Title : 8080.M
 QLast Update : Thu Feb 28 03:00:19 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	11.260	10.764	220.3E6	207.1E6	501.781	441.099
Target Compounds							
1) T	Dalapon	2.835	2.296	197.2E6	235.4E6	303.456	371.787
2) T	3,5-DICHL...	9.691	8.927	227.0E6	209.0E6	340.537	338.558m
3) T	4-Nitroph...	10.878	9.926	24943729	29379122	118.549m	130.674m
5) T	DICAMBA	11.564	11.042	618.2E6	585.5E6	350.703	350.482
6) T	MCPD	11.905	11.270	49517102	38144075	37.716m	32.314m
7) T	MCPA	12.154	11.645	68773622	55417846	34.672m	32.128
8) T	DICHLORPROP	12.757	12.168	152.0E6	145.5E6	351.217	352.003
9) T	2,4-D	13.122	12.642	156.8E6	184.6E6	337.975m	351.778
10) T	Pentachlo...	13.482	13.172	2262.5E6	2014.1E6	320.262	305.830m
11) T	2,4,5-TP ...	14.359	13.798	890.4E6	850.3E6	331.448	333.929
12) T	2,4,5-T	14.762	14.364	851.6E6	485.6E6	322.368	240.772 #
13) T	2,4-DB	15.508	15.015	99488860	104.7E6	420.146m	409.961m
14) T	DINOSEB	16.969	15.437	85774121	50668787	38.203m	28.388m#
15) T	Picloram	16.751	16.669	1023.9E6	1019.3E6	416.024m	401.591
16) T	DCPA	17.314	16.612	608.8E6	665.3E6	184.749	213.893

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS032519\
Data File : PS003722.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 25 Mar 2019 17:35
Operator : EI\SJ
Sample : K2002-01MS
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
S-1(2-10)-COMPMS

Manual Integrations
APPROVED

Sohil
3/26/2019 1:43:44 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 26 05:11:09 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS022719.M
Quant Title : 8080.M
QLast Update : Thu Feb 28 03:00:19 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

