

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\
 Method File : PS032619.M
 Title : 8080.M
 Last Update : Wed Mar 27 05:52:44 2019
 Response Via : Initial Calibration

Calibration Files

750 =PS003749.D 1000 =PS003750.D 1500 =PS003751.D
 500 =PS003748.D 200 =PS003747.D

	Compound	750	1000	1500	500	200	Avg		%RSD
1) T	Dalapon	7.607	7.371	7.241	7.534	7.524	7.455	E5	1.98
2) T	3,5-DICHLOROBENZOI...	6.979	6.682	6.655	6.861	7.191	6.874	E5	3.22
3) T	4-Nitrophenol	2.338	2.212	2.308	2.134	1.842	2.167	E5	9.18
4) S	2,4-DCAA	4.464	4.219	4.262	4.331	4.535	4.362	E5	3.07
5) T	DICAMBA	2.101	2.026	2.009	2.053	2.069	2.051	E6	1.76
6) T	MCPP	1.443	1.438	1.420	1.397	1.257	1.391	E6	5.52
7) T	MCPA	2.169	2.131	2.146	2.092	2.010	2.110	E6	2.96
8) T	DICHLORPROP	4.989	4.648	4.724	4.857	5.204	4.884	E5	4.52
9) T	2,4-D	5.057	4.926	4.961	4.825	4.545	4.863	E5	4.03
10) T	Pentachlorophenol	7.874	7.459	7.073	7.936	8.174	7.703	E6	5.66
11) T	2,4,5-TP (SILVEX)	2.932	2.824	2.750	2.905	2.932	2.869	E6	2.78
12) T	2,4,5-T	2.921	2.796	2.745	2.856	2.884	2.840	E6	2.46
13) T	2,4-DB	3.146	2.967	3.134	2.982	2.845	3.015	E5	4.18
14) T	DINOSEB	2.404	2.277	2.210	2.423	2.643	2.391	E6	6.94
15) T	Picloram	3.577	3.530	3.549	3.332	2.998	3.397	E6	7.15
16) T	DCPA	3.671	3.488	3.365	3.671	3.810	3.601	E6	4.85

Signal #2 Calibration Files

750 =PS003749.D 1000 =PS003750.D 1500 =PS003751.D
 500 =PS003748.D 200 =PS003747.D

	Compound	750	1000	1500	500	200	Avg		%RSD
1) T	Dalapon	7.292	7.045	6.850	7.328	7.966	7.296	E5	5.78
2) T	3,5-DICHLOROBENZOI...	6.420	6.187	6.044	6.407	6.835	6.379	E5	4.70
3) T	4-Nitrophenol	2.327	2.214	2.190	2.299	2.321	2.270	E5	2.80
4) S	2,4-DCAA	4.994	4.686	4.761	4.899	5.194	4.907	E5	4.08
5) T	DICAMBA	1.903	1.827	1.800	1.865	1.903	1.860	E6	2.46
6) T	MCPP	1.222	1.207	1.210	1.169	1.096	1.181	E6	4.35
7) T	MCPA	1.843	1.779	1.805	1.763	1.727	1.783	E6	2.45
8) T	DICHLORPROP	4.692	4.498	4.451	4.651	4.986	4.655	E5	4.52
9) T	2,4-D	5.393	5.143	5.146	5.281	5.632	5.319	E5	3.83
10) T	Pentachlorophenol	7.051	6.659	6.214	7.164	7.513	6.920	E6	7.21
11) T	2,4,5-TP (SILVEX)	2.687	2.585	2.500	2.684	2.785	2.648	E6	4.12
12) T	2,4,5-T	2.194	2.111	1.993	2.192	2.379	2.174	E6	6.49
13) T	2,4-DB	3.010	2.951	2.938	2.970	2.978	2.970	E5	0.93
14) T	DINOSEB	1.948	1.876	1.836	1.930	2.003	1.919	E6	3.39
15) T	Picloram	3.314	3.225	3.200	3.136	2.981	3.171	E6	3.91
16) T	DCPA	3.439	3.281	3.141	3.452	3.587	3.380	E6	5.09

(#) = Out of Range