

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS032620\
 Data File : PS009435.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 26 Mar 2020 12:44
 Operator : DD\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Manual Integrations
 APPROVED

Sohil
 3/30/2020 11:31:18 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 26 13:40:11 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS031120.M
 Quant Title : 8080.M
 QLast Update : Thu Mar 12 07:13:24 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.242	6.932	457.9E6	408.0E6	827.787m	742.875
Target Compounds							
1) T	Dalapon	3.047	2.629	585.4E6	555.3E6	750.668	705.299m
2) T	3,5-DICHL...	6.451	6.078	595.7E6	555.1E6	789.177	682.553
3) T	4-Nitroph...	7.071	6.534	182.6E6	211.8E6	765.009	665.757
5) T	DICAMBA	7.435	7.102	1688.2E6	1548.7E6	793.251	674.647
6) T	MCPD	7.606	7.201	134.8E6	115.6E6	54.790	63.559m
7) T	MCPA	7.756	7.417	187.3E6	184.5E6	61.490	68.366
8) T	DICHLORPROP	8.129	7.729	404.5E6	388.8E6	784.215	700.741
9) T	2,4-D	8.357	8.023	418.0E6	456.9E6	792.799	695.106
10) T	Pentachlo...	8.671	8.457	5371.8E6	4612.5E6	763.995	607.715
11) T	2,4,5-TP ...	9.220	8.817	2173.5E6	2059.9E6	792.896	690.976
12) T	2,4,5-T	9.504	9.196	2006.7E6	2032.9E6	788.849	688.618
13) T	2,4-DB	10.058	9.704	271.2E6	272.6E6	783.960	718.038
14) T	DINOSEB	11.224	10.040	1595.5E6	1535.8E6	780.541	699.223
15) T	Picloram	11.037	11.020	2254.1E6	2615.3E6	742.723	642.243
16) T	DCPA	11.488	10.970	2665.2E6	2640.2E6	768.901	681.677

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS032620\
 Data File : PS009435.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 26 Mar 2020 12:44
 Operator : DD\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Manual Integrations
 APPROVED

Sohil
 3/30/2020 11:31:18 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 26 13:40:11 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS031120.M
 Quant Title : 8080.M
 QLast Update : Thu Mar 12 07:13:24 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

