

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS032720\
 Data File : PS009479.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 27 Mar 2020 21:03
 Operator : DD\AJ
 Sample : L2040-09
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 RO-11986

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 28 05:48:10 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS031120.M
 Quant Title : 8080.M
 QLast Update : Thu Mar 12 07:13:24 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.240	6.931	223.7E6	179.4E6	404.311	326.738
Target Compounds							
1) T	Dalapon	3.006	2.586	21044857	64268016	26.986	81.624 #
2) T	3,5-DICHL...	0.000	6.108	0	50275805	N.D.	61.817
5) T	DICAMBA	0.000	7.152	0	6236082	N.D.	2.717
6) T	MCP	0.000	7.218	0	115.5E6	N.D.	63.490
8) T	DICHLORPROP	0.000	7.749	0	33829508	N.D.	60.974
9) T	2,4-D	8.294	0.000	8556060	0	16.228	N.D. #
10) T	Pentachlo...	8.669	8.455	2326.2E6	2236.1E6	330.839	294.611
11) T	2,4,5-TP ...	0.000	8.797	0	385.5E6	N.D.	129.325
12) T	2,4,5-T	9.443	9.262	59331700	3060385	23.324	1.037 #
13) T	2,4-DB	0.000	9.698	0	7560202	N.D.	19.915
14) T	DINOSEB	0.000	10.083	0	9917431	N.D.	4.515



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS032720\
Data File : PS009479.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Mar 2020 21:03
Operator : DD\AJ
Sample : L2040-09
Misc :
ALS Vial : 18 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampled :
RO-11986

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 28 05:48:10 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS031120.M
Quant Title : 8080.M
QLast Update : Thu Mar 12 07:13:24 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Dalapon


R.T.: 3.006 min
Delta R.T.: -0.040 min
Response: 21044857
Conc: 26.99 ng/ml

Instrument :
ECD_S
ClientSampleId :
RO-11986




#1 Dalapon

R.T.: 2.586 min
Delta R.T.: -0.040 min
Response: 64268016
Conc: 81.62 ng/ml




#2 3,5-DICHLOROBENZOIC ACID

R.T.: 0.000 min
Exp R.T. : 6.451 min
Response: 0
Conc: N.D.



#2 3,5-DICHLOROBENZOIC ACID


R.T.: 6.108 min
Delta R.T.: 0.028 min
Response: 50275805
Conc: 61.82 ng/ml



#4 2,4-DCAA


R.T.: 7.240 min
Delta R.T.: -0.002 min
Response: 223652530
Conc: 404.31 ng/ml

Instrument :
ECD_S
ClientSampleId :
RO-11986




#4 2,4-DCAA

R.T.: 6.931 min
Delta R.T.: -0.004 min
Response: 179443798
Conc: 326.74 ng/ml




#5 DICAMBA

R.T.: 0.000 min
Exp R.T. : 7.435 min
Response: 0
Conc: N.D.



#5 DICAMBA


R.T.: 7.152 min
Delta R.T.: 0.047 min
Response: 6236082
Conc: 2.72 ng/ml



#6 MCP


R.T.: 0.000 min
Exp R.T.: 7.606 min
Response: 0
Conc: N.D.

Instrument :
ECD_S
ClientSampleId :
RO-11986




#6 MCP

R.T.: 7.218 min
Delta R.T.: 0.015 min
Response: 115476511
Conc: 63.49 ug/ml




#8 DICHLORPROP

R.T.: 0.000 min
Exp R.T.: 8.129 min
Response: 0
Conc: N.D.



#8 DICHLORPROP


R.T.: 7.749 min
Delta R.T.: 0.016 min
Response: 33829508
Conc: 60.97 ng/ml



#9 2,4-D


R.T.: 8.294 min
Delta R.T.: -0.064 min
Response: 8556060
Conc: 16.23 ng/ml

Instrument :
ECD_S
ClientSampleId :
RO-11986




#9 2,4-D

R.T.: 0.000 min
Exp R.T. : 8.027 min
Response: 0
Conc: N.D.




#10 Pentachlorophenol

R.T.: 8.669 min
Delta R.T.: -0.003 min
Response: 2326206035
Conc: 330.84 ng/ml



#10 Pentachlorophenol


R.T.: 8.455 min
Delta R.T.: -0.006 min
Response: 2236091559
Conc: 294.61 ng/ml



#11 2,4,5-TP (SILVEX)


R.T.: 0.000 min
Exp R.T.: 9.221 min
Response: 0
Conc: N.D.

Instrument :
ECD_S
ClientSampleId :
RO-11986




#11 2,4,5-TP (SILVEX)

R.T.: 8.797 min
Delta R.T.: -0.024 min
Response: 385533796
Conc: 129.33 ng/ml




#12 2,4,5-T

R.T.: 9.443 min
Delta R.T.: -0.062 min
Response: 59331700
Conc: 23.32 ng/ml



#12 2,4,5-T


R.T.: 9.262 min
Delta R.T.: 0.061 min
Response: 3060385
Conc: 1.04 ng/ml



#13 2,4-DB


R.T.: 0.000 min
Exp R.T.: 10.059 min
Response: 0
Conc: N.D.

Instrument :
ECD_S
ClientSampleId :
RO-11986




#13 2,4-DB

R.T.: 9.698 min
Delta R.T.: -0.010 min
Response: 7560202
Conc: 19.92 ng/ml



#14 DINOSEB

R.T.: 0.000 min
Exp R.T.: 11.225 min
Response: 0
Conc: N.D.



#14 DINOSEB

R.T.: 10.083 min
Delta R.T.: 0.040 min
Response: 9917431
Conc: 4.52 ng/ml