

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS032723\
 Data File : PS022526.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 27 Mar 2023 17:56
 Operator : AR\AJ
 Sample : 02047-01MS
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 R-PITMS

Manual Integrations
 APPROVED

Reviewed By : Abdul Mirza 03/28/2023
 Supervised By : Ankita Jodhani 03/28/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 27 21:17:21 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS032023.M
 Quant Title : 8080.M
 QLast Update : Mon Mar 20 15:46:44 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	6.292	6.716	622.5E6	600.6E6	254.673m	291.616
Target Compounds							
2) T	3,5-DICHL...	5.677	5.920	727.5E6	583.3E6	195.395	182.907
5) T	DICAMBA	6.430	6.874	2167.1E6	1979.1E6	210.799	200.616
6) T	MCP	6.576	6.966	208.2E6	177.3E6	27.386m	22.775
7) T	MCPA	6.699	7.175	209.0E6	283.0E6	20.027	26.939m#
8) T	DICHLORPROP	7.013	7.495	495.8E6	535.3E6	187.064m	214.997
9) T	2,4-D	7.215	7.784	706.5E6	635.0E6	215.472	214.624
10) T	Pentachlo...	7.435	8.203	2064.3E6	1739.1E6	58.670	52.983
11) T	2,4,5-TP ...	7.961	8.566	2022.5E6	2095.2E6	135.308	153.815
12) T	2,4,5-T	8.220	8.942	2785.7E6	2246.6E6	169.639	178.589
13) T	2,4-DB	8.729	9.445	490.4E6	410.5E6	162.625	244.442m#
15) T	Picloram	9.610	10.739	5564.3E6	4561.4E6	224.173	227.307
16) T	DCPA	10.035	10.693	6246.3E6	5142.3E6	302.832m	324.185

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS032723\
Data File : PS022526.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Mar 2023 17:56
Operator : AR\AJ
Sample : 02047-01MS
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
R-PITMS

Manual Integrations
APPROVED

Reviewed By : Abdul Mirza 03/28/2023
Supervised By : Ankita Jodhani 03/28/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 27 21:17:21 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS032023.M
Quant Title : 8080.M
QLast Update : Mon Mar 20 15:46:44 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

