

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS032725\
 Data File : PS029545.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 26 Mar 2025 13:04
 Operator : AR\AJ
 Sample : HSTDICC500
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 26 13:25:14 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS032625.M
 Quant Title : 8080.M
 QLast Update : Wed Mar 26 04:43:02 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.126	7.542	1934.1E6	428.2E6	500.465	506.212
Target Compounds							
1) T	Dalapon	2.593	2.609	3033.6E6	1012.9E6	538.210	505.584
2) T	3,5-DICHL...	6.314	6.530	2589.6E6	586.3E6	480.228	471.679
3) T	4-Nitroph...	6.956	7.080	225.8E6	377.7E6	253.676	463.229 #
5) T	DICAMBA	7.302	7.731	8408.6E6	2235.0E6	567.349	479.415
6) T	MCPD	7.478	7.834	515.5E6	102.0E6	48.152	52.378
7) T	MCPA	7.624	8.070	606.7E6	136.0E6	44.612	48.874
8) T	DICHLORPROP	8.000	8.434	1908.4E6	536.6E6	422.474	418.451
9) T	2,4-D	8.240	8.756	1581.9E6	593.5E6	424.858	406.578
10) T	Pentachlo...	8.516	9.262	30367.7E6	10777.8E6	493.889	491.362
11) T	2,4,5-TP ...	9.091	9.640	10478.1E6	4133.9E6	473.336	467.802
12) T	2,4,5-T	9.390	10.051	7825.1E6	3789.3E6	459.865	499.875
13) T	2,4-DB	9.966	10.610	440.9E6	385.3E6	343.122	482.426 #
14) T	DINOSEB	11.135	10.981	7749.3E6	2789.0E6	473.848	465.034
15) T	Picloram	10.990	12.046	2634.8E6	5712.8E6	286.694	472.953 #
16) T	DCPA	11.432	12.008	13063.7E6	5464.9E6	501.740	496.166

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS032725\
Data File : PS029545.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 26 Mar 2025 13:04
Operator : AR\AJ
Sample : HSTDICC500
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 26 13:25:14 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS032625.M
Quant Title : 8080.M
QLast Update : Wed Mar 26 04:43:02 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

