

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS032822\
 Data File : PS018790.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 28 Mar 2022 14:38
 Operator : AJ\MA
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 28 17:41:54 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS031422.M
 Quant Title : 8080.M
 QLast Update : Mon Mar 14 11:41:05 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.228	6.809	2141.7E6	518.6E6	752.552	736.991
Target Compounds							
1) T	Dalapon	3.032	2.536	2302.2E6	556.1E6	592.066	672.843
2) T	3,5-DICHL...	6.437	5.982	2675.9E6	726.8E6	694.389	694.726
3) T	4-Nitroph...	7.055	6.420	1035.9E6	269.1E6	601.162	662.341
5) T	DICAMBA	7.421	6.976	7547.9E6	2067.0E6	692.043	693.960
6) T	MCPD	7.592	7.075	594.5E6	163.4E6	70.034	66.756
7) T	MCPA	7.742	7.289	835.1E6	236.2E6	69.048	67.648
8) T	DICHLORPROP	8.116	7.600	2022.2E6	550.6E6	734.157	695.490
9) T	2,4-D	8.344	7.891	2437.8E6	655.4E6	727.579	700.202
10) T	Pentachlo...	8.655	8.321	27396.8E6	6993.4E6	756.762	712.133
11) T	2,4,5-TP ...	9.208	8.682	11842.3E6	3135.5E6	778.677	718.827
12) T	2,4,5-T	9.492	9.059	11782.3E6	2985.7E6	788.709	720.813
13) T	2,4-DB	10.044	9.566	2179.1E6	387.3E6	796.637	682.084
14) T	DINOSEB	11.210	9.895	8668.9E6	2129.3E6	810.697	694.963
15) T	Picloram	11.026	10.874	17199.2E6	3953.9E6	836.711	743.755
16) T	DCPA	11.476	10.828	15512.9E6	3899.9E6	867.087	741.527

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS032822\
Data File : PS018790.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 28 Mar 2022 14:38
Operator : AJ\MA
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 28 17:41:54 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS031422.M
Quant Title : 8080.M
QLast Update : Mon Mar 14 11:41:05 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

