

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS032822\
 Data File : PS018805.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 28 Mar 2022 22:45
 Operator : AJ\MA
 Sample : N2125-03MSD
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 164718MSD

Manual Integrations
 APPROVED

Reviewed By :Abdul Mirza 03/30/2022
 Supervised By :mohammad ahmed 03/30/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 29 01:12:57 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS031422.M
 Quant Title : 8080.M
 QLast Update : Mon Mar 14 11:41:05 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.226	6.811	1298.2E6	375.8E6	456.157m	534.004m
Target Compounds							
1) T	Dalapon	3.029	2.540	645.9E6	478.3E6	166.107	578.729m#
2) T	3,5-DICHL...	6.435	5.983	1437.8E6	317.4E6	373.107m	303.439m
3) T	4-Nitroph...	7.055	6.419	556.9E6	225.2E6	323.165	554.257 #
5) T	DICAMBA	7.420	6.979	3510.0E6	983.2E6	321.823	330.084
6) T	MCPD	7.589	7.082	274.7E6	67225380	32.361	27.460
7) T	MCPA	7.743	7.286	359.8E6	111.6E6	29.749	31.960
8) T	DICHLORPROP	8.116	7.640	1043.0E6	4375.5E6	378.681	5527.325 #
9) T	2,4-D	8.348	7.894	1380.8E6	366.9E6	412.103	391.976
10) T	Pentachlo...	8.657	8.324	12869.3E6	3423.9E6	355.479	348.652
11) T	2,4,5-TP ...	9.212	8.684	5824.1E6	1589.6E6	382.956	364.416
12) T	2,4,5-T	9.496	9.114f	5230.9E6	3948.8E6	350.158	953.307m#
13) T	2,4-DB	10.039	9.558	3451.0E6	74431.6E6	1261.620	131099.618 #
15) T	Picloram	11.048	10.887	6308.6E6	2012.0E6	306.903m	378.477
16) T	DCPA	11.481	10.836	7768.3E6	2798.2E6	434.209m	532.050

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS032822\
Data File : PS018805.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 28 Mar 2022 22:45
Operator : AJ\MA
Sample : N2125-03MSD
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
164718MSD

Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 03/30/2022
Supervised By :mohammad ahmed 03/30/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 29 01:12:57 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS031422.M
Quant Title : 8080.M
QLast Update : Mon Mar 14 11:41:05 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

