

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS032825\
 Data File : PS029565.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 27 Mar 2025 13:13
 Operator : AR\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 27 13:42:57 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS032825.M
 Quant Title : 8080.M
 QLast Update : Thu Mar 27 12:24:58 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.115	7.535	2828.2E6	638.7E6	736.481	739.822
Target Compounds							
1) T	Dalapon	2.594	2.610	4131.3E6	1390.8E6	696.860	670.484
2) T	3,5-DICHL...	6.304	6.524	3939.0E6	855.2E6	689.056	694.053
3) T	4-Nitroph...	6.923	7.073	742.3E6	547.4E6	697.642	677.559
5) T	DICAMBA	7.294	7.725	11009.2E6	3412.8E6	694.458	713.861
6) T	MCPD	7.474	7.832	825.6E6	151.2E6	72.528	73.046
7) T	MCPA	7.621	8.067	1023.1E6	203.0E6	71.135	70.623
8) T	DICHLORPROP	7.991	8.427	3037.4E6	863.4E6	691.588	693.138
9) T	2,4-D	8.222	8.747	2783.9E6	959.9E6	690.142	702.681
10) T	Pentachlo...	8.511	9.256	44203.8E6	16102.5E6	743.921	725.510
11) T	2,4,5-TP ...	9.082	9.633	16479.5E6	6377.8E6	707.424	722.722
12) T	2,4,5-T	9.374	10.042	14543.3E6	5752.5E6	706.043	721.706
13) T	2,4-DB	9.944	10.601	1333.0E6	611.0E6	698.158	664.909
14) T	DINOSEB	11.127	10.974	12329.4E6	4225.1E6	695.392	712.747
15) T	Picloram	10.954	12.035	9871.2E6	9061.4E6	685.813	726.299
16) T	DCPA	11.424	12.000	20561.1E6	8222.8E6	713.160	733.404

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS032825\
Data File : PS029565.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Mar 2025 13:13
Operator : AR\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 27 13:42:57 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS032825.M
Quant Title : 8080.M
QLast Update : Thu Mar 27 12:24:58 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

