

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS032919\
 Data File : PS003837.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 29 Mar 2019 20:01
 Operator : EI\SJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Manual Integrations
 APPROVED

Sohil
 3/30/2019 1:11:15 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 30 02:24:49 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS032619.M
 Quant Title : 8080.M
 QLast Update : Wed Mar 27 05:52:44 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	11.286	10.787	308.6E6	355.0E6	707.456	723.516
Target Compounds							
1) T	Dalapon	2.854	2.328	506.9E6	485.1E6	679.886	664.890
2) T	3,5-DICHL...	9.714	8.950	464.0E6	435.9E6	675.112	683.373
3) T	4-Nitroph...	10.897	9.946	133.4E6	146.8E6	615.566	646.487
5) T	DICAMBA	11.586	11.065	1435.9E6	1300.5E6	699.962	699.296
6) T	MCPD	11.928	11.295	99886726	82477937	71.808m	69.854
7) T	MCPA	12.178	11.671	145.3E6	119.8E6	68.860	67.183m
8) T	DICHLORPROP	12.779	12.190	327.8E6	318.5E6	671.025	684.069
9) T	2,4-D	13.148	12.666	317.1E6	352.5E6	652.044	662.791
10) T	Pentachlo...	13.502	13.193	5467.5E6	4918.3E6	709.757	710.700
11) T	2,4,5-TP ...	14.381	13.820	2015.7E6	1868.5E6	702.668	705.537
12) T	2,4,5-T	14.784	14.385	1959.3E6	1488.9E6	689.803	684.919m
13) T	2,4-DB	15.532	15.038	187.6E6	190.0E6	622.132	640.007
14) T	DINOSEB	16.988	15.456	1691.4E6	1305.8E6	707.300	680.554
15) T	Picloram	16.778	16.692	1994.5E6	1894.6E6	587.136	597.433
16) T	DCPA	17.335	16.630	2552.9E6	2380.1E6	708.971	704.164

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS032919\
Data File : PS003837.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 29 Mar 2019 20:01
Operator : EI\SJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Manual Integrations
APPROVED

Sohil
3/30/2019 1:11:15 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 30 02:24:49 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS032619.M
Quant Title : 8080.M
QLast Update : Wed Mar 27 05:52:44 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

