

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS032921\
 Data File : PS014396.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 30 Mar 2021 03:05
 Operator : DD\AJ
 Sample : PB135400BS
 Misc :
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 PB135400BS

Manual Integrations
 APPROVED

Ankita
 3/30/2021 11:43:41 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 30 06:20:17 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS031721.M
 Quant Title : 8080.M
 QLast Update : Thu Mar 18 02:23:19 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.164	6.888	832.3E6	292.5E6	582.520	620.066m
Target Compounds							
1) T	Dalapon	3.008	2.605	848.3E6	344.4E6	545.013	581.636
2) T	3,5-DICHL...	6.388	6.049	1087.5E6	391.5E6	563.316	525.832m
3) T	4-Nitroph...	6.998	6.504	397.4E6	142.1E6	532.102	467.949
5) T	DICAMBA	7.352	7.054	2914.5E6	1117.4E6	541.953	557.777
6) T	MCPD	7.517	7.148	249.0E6	96002321	59.892	57.655
7) T	MCPA	7.665	7.362	299.2E6	150.7E6	52.002	61.849
8) T	DICHLORPROP	8.038	7.679	737.3E6	289.2E6	556.768	567.632
9) T	2,4-D	8.267	7.976	883.8E6	365.5E6	599.634	612.202
10) T	Pentachlo...	8.572	8.399	11559.1E6	3978.7E6	580.723	582.571
11) T	2,4,5-TP ...	9.117	8.760	4588.9E6	1676.3E6	595.546	616.148
12) T	2,4,5-T	9.402	9.141	4126.8E6	1444.8E6	530.371	541.920
13) T	2,4-DB	9.953	9.646	506.5E6	183.6E6	327.561	517.847 #
14) T	DINOSEB	11.103	9.975	3219.4E6	1012.3E6	563.255	560.173
15) T	Picloram	10.929	10.960	3804.4E6	1415.0E6	364.319	372.982
16) T	DCPA	11.369	10.901	5255.6E6	2013.3E6	446.212	490.812

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS032921\
 Data File : PS014396.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 30 Mar 2021 03:05
 Operator : DD\AJ
 Sample : PB135400BS
 Misc :
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 PB135400BS

Manual Integrations
 APPROVED

Ankita
 3/30/2021 11:43:41 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 30 06:20:17 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS031721.M
 Quant Title : 8080.M
 QLast Update : Thu Mar 18 02:23:19 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

