

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS032923\
 Data File : PS022546.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 29 Mar 2023 16:12
 Operator : AR\AJ
 Sample : 02102-02MS
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 CHAMBER-C-72-EASTMS

Manual Integrations
 APPROVED

Reviewed By : Abdul Mirza 03/30/2023
 Supervised By : Ankita Jodhani 03/30/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 29 22:02:34 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS032023.M
 Quant Title : 8080.M
 QLast Update : Mon Mar 20 15:46:44 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	6.295	6.718	1137.1E6	1198.7E6	465.199m	581.971 #
Target Compounds							
1) T	Dalapon	2.408	2.481	867.4E6	1709.8E6	278.129m	523.906m#
2) T	3,5-DICHL...	5.677	5.920	1655.9E6	1486.7E6	444.738	466.193
5) T	DICAMBA	6.432	6.876	4386.9E6	4285.9E6	426.729	434.459
6) T	MCPPE	6.581	6.972	353.0E6	360.5E6	46.436	46.315
7) T	MCPA	6.702	7.181	417.5E6	455.7E6	40.001	43.372
8) T	DICHLORPROP	7.015	7.496	1174.0E6	1381.6E6	442.984	554.864 #
9) T	2,4-D	7.217	7.785	1466.2E6	1492.8E6	447.146	504.582
10) T	Pentachlo...	7.436	8.203	17886.3E6	16946.0E6	508.347	516.263
11) T	2,4,5-TP ...	7.962	8.566	7345.1E6	6716.9E6	491.390	493.099
12) T	2,4,5-T	8.221	8.942	7522.6E6	6268.0E6	458.104	498.261
13) T	2,4-DB	8.728	9.447	1028.8E6	781.5E6	341.192	465.355m#
14) T	DINOSEB	9.768	9.769	5839.0E6	4499.1E6	450.552	469.085
15) T	Picloram	9.616	10.743	9407.3E6	8137.7E6	379.002	405.526
16) T	DCPA	10.036	10.693	10144.4E6	7800.6E6	491.817	491.771

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS032923\
Data File : PS022546.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 29 Mar 2023 16:12
Operator : AR\AJ
Sample : 02102-02MS
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
CHAMBER-C-72-EASTMS

Manual Integrations
APPROVED

Reviewed By : Abdul Mirza 03/30/2023
Supervised By : Ankita Jodhani 03/30/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 29 22:02:34 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS032023.M
Quant Title : 8080.M
QLast Update : Mon Mar 20 15:46:44 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

