

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS033021\
 Data File : PS014402.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 30 Mar 2021 08:45
 Operator : DD\AJ
 Sample : HEXANE
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 30 11:19:59 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS031721.M
 Quant Title : 8080.M
 QLast Update : Thu Mar 18 02:23:19 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
----------	------	------	--------	--------	-------	-------

System Monitoring Compounds

Target Compounds

6) T	MCP	7.567f	7.188f	922490	220853	<MDL	<MDL	#
7) T	MCPA	7.623f	0.000	10995674	0	1.911	N.D.	#
9) T	2,4-D	8.330f	0.000	14172507	0	9.615	N.D.	#
12) T	2,4,5-T	9.362f	0.000	22529472	0	2.895	N.D.	#
15) T	Picloram	10.924	11.006f	3625713	11215954	<MDL	2.956	#
16) T	DCPA	11.372	0.000	5470591	0	<MDL	N.D.	#



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS033021\
Data File : PS014402.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 30 Mar 2021 08:45
Operator : DD\AJ
Sample : HEXANE
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 30 11:19:59 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS031721.M
Quant Title : 8080.M
QLast Update : Thu Mar 18 02:23:19 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#4 2,4-DCAA


R.T.: 0.000 min
Exp R.T.: 7.170 min
Response: 0
Conc: N.D.

Instrument :
ECD_S
ClientSampleId :




#4 2,4-DCAA

R.T.: 0.000 min
Exp R.T.: 6.892 min
Response: 0
Conc: N.D.




#7 MCPA

R.T.: 7.623 min
Delta R.T.: -0.043 min
Response: 10995674
Conc: 1.91 ug/ml



#7 MCPA


R.T.: 0.000 min
Exp R.T.: 7.359 min
Response: 0
Conc: N.D.



#9 2,4-D


R.T.: 8.330 min
Delta R.T.: 0.056 min
Response: 14172507
Conc: 9.62 ng/ml

Instrument :
ECD_S
ClientSampleId :




#9 2,4-D

R.T.: 0.000 min
Exp R.T. : 7.982 min
Response: 0
Conc: N.D.




#12 2,4,5-T

R.T.: 9.362 min
Delta R.T.: -0.048 min
Response: 22529472
Conc: 2.90 ng/ml



#12 2,4,5-T


R.T.: 0.000 min
Exp R.T. : 9.145 min
Response: 0
Conc: N.D.



#15 Picloram

R.T.: 10.924 min
Delta R.T.: -0.015 min
Response: 3625713
Conc: N.D.

Instrument :
ECD_S
ClientSampleId :



#15 Picloram

R.T.: 11.006 min
Delta R.T.: 0.042 min
Response: 11215954
Conc: 2.96 ng/ml