

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS033021\
 Data File : PS014405.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 30 Mar 2021 13:34
 Operator : DD\AJ
 Sample : PB135413BS
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 PB135413BS

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 30 14:22:33 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS031721.M
 Quant Title : 8080.M
 QLast Update : Thu Mar 18 02:23:19 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.160	6.885	732.0E6	218.0E6	512.283	462.168
Target Compounds							
1) T	Dalapon	3.004	2.600	632.0E6	221.4E6	406.025	373.879
2) T	3,5-DICHL...	6.385	6.047	913.5E6	320.2E6	473.159	430.115
3) T	4-Nitroph...	6.994	6.502	341.8E6	119.8E6	457.718	394.351
5) T	DICAMBA	7.348	7.051	2674.7E6	929.7E6	497.354	464.093
6) T	MCPD	7.512	7.143	187.4E6	72109063	45.068	43.306
7) T	MCPA	7.660	7.358	260.6E6	101.1E6	45.297	41.498
8) T	DICHLORPROP	8.034	7.677	651.4E6	232.7E6	491.910	456.793
9) T	2,4-D	8.264	7.973	763.7E6	262.5E6	518.099	439.641
10) T	Pentachlo...	8.569	8.396	10315.1E6	3330.9E6	518.229	487.709
11) T	2,4,5-TP ...	9.114	8.758	3871.9E6	1285.1E6	502.483	472.334
12) T	2,4,5-T	9.400	9.138	3736.8E6	1184.9E6	480.256	444.440
13) T	2,4-DB	9.949	9.645	574.8E6	149.2E6	371.695	420.679
14) T	DINOSEB	11.100	9.973	2913.6E6	865.1E6	509.759	478.703
15) T	Picloram	10.927	10.959	3673.8E6	1211.9E6	351.820	319.444
16) T	DCPA	11.366	10.900	5146.4E6	1717.1E6	436.939	418.609

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS033021\
Data File : PS014405.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 30 Mar 2021 13:34
Operator : DD\AJ
Sample : PB135413BS
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
PB135413BS

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 30 14:22:33 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS031721.M
Quant Title : 8080.M
QLast Update : Thu Mar 18 02:23:19 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

