

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS033121\
 Data File : PS014431.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 31 Mar 2021 20:33
 Operator : DD\AJ
 Sample : I.BLK
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 I.BLK

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 01 02:01:27 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS031721.M
 Quant Title : 8080.M
 QLast Update : Thu Mar 18 02:23:19 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.161	6.887	674.4E6	259.1E6	472.008	549.270
Target Compounds							
5) T	DICAMBA	0.000	7.008f	0	710442	N.D.	<MDL
9) T	2,4-D	8.332f	0.000	16707945	0	11.335	N.D. #
13) T	2,4-DB	9.985	0.000	10306236	0	6.665	N.D. #
15) T	Picloram	0.000	10.988	0	4846598	N.D.	1.278
16) T	DCPA	0.000	10.880	0	2886424	N.D.	<MDL



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS033121\
Data File : PS014431.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 31 Mar 2021 20:33
Operator : DD\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 01 02:01:27 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS031721.M
Quant Title : 8080.M
QLast Update : Thu Mar 18 02:23:19 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#4 2,4-DCAA


R.T.: 7.161 min
Delta R.T.: -0.009 min
Response: 674425413
Conc: 472.01 ng/ml

Instrument :
ECD_S
ClientSampleId :
I.BLK




#4 2,4-DCAA

R.T.: 6.887 min
Delta R.T.: -0.005 min
Response: 259128615
Conc: 549.27 ng/ml




#9 2,4-D

R.T.: 8.332 min
Delta R.T.: 0.058 min
Response: 16707945
Conc: 11.34 ng/ml



#9 2,4-D


R.T.: 0.000 min
Exp R.T.: 7.982 min
Response: 0
Conc: N.D.



#13 2,4-DB


R.T.: 9.985 min
Delta R.T.: 0.025 min
Response: 10306236
Conc: 6.66 ng/ml

Instrument :
ECD_S
ClientSampleId :
I.BLK




#13 2,4-DB

R.T.: 0.000 min
Exp R.T. : 9.651 min
Response: 0
Conc: N.D.



#15 Picloram

R.T.: 0.000 min
Exp R.T. : 10.939 min
Response: 0
Conc: N.D.



#15 Picloram

R.T.: 10.988 min
Delta R.T.: 0.024 min
Response: 4846598
Conc: 1.28 ng/ml