

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS040119\
 Data File : PS003891.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 01 Apr 2019 17:55
 Operator : EI\SJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Manual Integrations
 APPROVED

Sohil
 4/11/2019 3:41:48 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 02 05:20:56 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS032619.M
 Quant Title : 8080.M
 QLast Update : Wed Mar 27 05:52:44 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	11.282	10.783	313.6E6	361.2E6	718.899	736.063
Target Compounds							
1) T	Dalapon	2.851	2.325	517.7E6	492.4E6	694.380	674.834
2) T	3,5-DICHL...	9.710	8.945	473.5E6	443.6E6	688.928	695.420
3) T	4-Nitroph...	10.899	9.943	127.2E6	145.7E6	587.145	641.776
5) T	DICAMBA	11.581	11.060	1462.6E6	1321.1E6	712.959	710.339
6) T	MCPD	11.923	11.291	102.4E6	83285806	73.642m	70.539
7) T	MCPA	12.173	11.666	149.2E6	123.4E6	70.709	69.205
8) T	DICHLORPROP	12.775	12.186	335.7E6	324.3E6	687.335	696.530
9) T	2,4-D	13.144	12.662	315.3E6	352.0E6	648.454	661.780
10) T	Pentachlo...	13.497	13.189	5561.5E6	4962.4E6	721.958	717.066
11) T	2,4,5-TP ...	14.376	13.816	2053.7E6	1894.4E6	715.921	715.317
12) T	2,4,5-T	14.780	14.381	1986.8E6	1413.3E6	699.470	650.121m
13) T	2,4-DB	15.528	15.033	189.8E6	193.3E6	629.613	651.088
14) T	DINOSEB	16.983	15.451	1731.3E6	1329.6E6	723.969	692.989
15) T	Picloram	16.772	16.687	2129.8E6	1954.9E6	626.947m	616.447
16) T	DCPA	17.329	16.625	2607.9E6	2415.4E6	724.224	714.611

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS040119\
Data File : PS003891.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 01 Apr 2019 17:55
Operator : EI\SJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampled :
HSTDCCC750

Manual Integrations
APPROVED

Sohil
4/11/2019 3:41:48 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 02 05:20:56 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS032619.M
Quant Title : 8080.M
QLast Update : Wed Mar 27 05:52:44 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

