

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS040121\
 Data File : PS014435.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 01 Apr 2021 13:20
 Operator : DD\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 01 14:08:50 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS031721.M
 Quant Title : 8080.M
 QLast Update : Thu Mar 18 02:23:19 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.158	6.884	1111.3E6	343.9E6	777.747	728.903
Target Compounds							
1) T	Dalapon	3.004	2.601	990.8E6	364.3E6	636.529	615.254
2) T	3,5-DICHL...	6.384	6.045	1381.7E6	496.4E6	715.725	666.646
3) T	4-Nitroph...	6.992	6.498	574.0E6	203.2E6	768.624	669.093
5) T	DICAMBA	7.347	7.051	3988.6E6	1399.4E6	741.678	698.527
6) T	MCPD	7.515	7.146	297.7E6	113.1E6	71.600	67.936
7) T	MCPA	7.663	7.361	417.9E6	161.6E6	72.646	66.331
8) T	DICHLORPROP	8.033	7.675	959.9E6	346.5E6	724.866	680.123
9) T	2,4-D	8.263	7.971	1094.9E6	391.4E6	742.824	655.577
10) T	Pentachlo...	8.569	8.396	14536.9E6	4700.8E6	730.327	688.294
11) T	2,4,5-TP ...	9.114	8.757	5710.3E6	1905.1E6	741.075	700.225
12) T	2,4,5-T	9.398	9.136	5695.9E6	1858.4E6	732.042	697.068
13) T	2,4-DB	9.945	9.642	1098.6E6	237.1E6	710.455	668.495
14) T	DINOSEB	11.101	9.972	4345.7E6	1326.6E6	760.305	734.134
15) T	Picloram	10.924	10.955	7442.9E6	2510.1E6	712.758	661.654
16) T	DCPA	11.366	10.899	8691.3E6	2907.7E6	737.910	708.865

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS040121\
Data File : PS014435.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 01 Apr 2021 13:20
Operator : DD\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 01 14:08:50 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS031721.M
Quant Title : 8080.M
QLast Update : Thu Mar 18 02:23:19 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

