

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS040122\
 Data File : PS018876.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 01 Apr 2022 21:43
 Operator : AJ\MA
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Manual Integrations
 APPROVED

Reviewed By : Abdul Mirza 04/04/2022
 Supervised By : mohammad ahmed 04/04/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 01 23:27:32 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS033122.M
 Quant Title : 8080.M
 QLast Update : Thu Mar 31 16:06:48 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.226	6.808	2462.9E6	548.9E6	843.613m	768.839
Target Compounds							
1) T	Dalapon	3.029	2.533	3020.8E6	603.5E6	660.368	650.625
2) T	3,5-DICHL...	6.436	5.981	3281.2E6	786.9E6	817.915	713.623
3) T	4-Nitroph...	7.058	6.421	915.6E6	296.9E6	802.129m	698.983
5) T	DICAMBA	7.418	6.974	10216.8E6	2301.2E6	812.598m	722.862
6) T	MCP	7.589	7.073	857.0E6	184.2E6	83.999	71.986
7) T	MCPA	7.740	7.287	1157.6E6	256.5E6	81.323	71.592
8) T	DICHLORPROP	8.113	7.597	2628.7E6	593.4E6	788.562	721.543
9) T	2,4-D	8.343	7.889	3112.8E6	689.8E6	778.936	746.395
10) T	Pentachlo...	8.652	8.317	34435.7E6	7640.4E6	769.658	726.059
11) T	2,4,5-TP ...	9.205	8.678	14360.9E6	3371.9E6	701.849	737.629
12) T	2,4,5-T	9.489	9.057	14220.9E6	3153.6E6	693.657	757.302
13) T	2,4-DB	10.043	9.563	2470.0E6	421.2E6	727.558	739.558
14) T	DINOSEB	11.205	9.891	9952.0E6	2355.9E6	609.567	730.553
15) T	Picloram	11.025	10.873	18178.3E6	4021.2E6	649.735	711.178
16) T	DCPA	11.472	10.823	15962.3E6	4060.7E6	634.182	744.993

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS040122\
Data File : PS018876.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 01 Apr 2022 21:43
Operator : AJ\MA
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 04/04/2022
Supervised By :mohammad ahmed 04/04/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 01 23:27:32 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS033122.M
Quant Title : 8080.M
QLast Update : Thu Mar 31 16:06:48 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

