

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS040125\
 Data File : PS029642.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 01 Apr 2025 15:35
 Operator : AR\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Manual Integrations
 APPROVED

Reviewed By : Abdul Mirza 04/02/2025
 Supervised By : mohammad ahmed 04/03/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 02 03:12:04 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS033125.M
 Quant Title : 8080.M
 QLast Update : Wed Apr 02 02:48:40 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	6.998	7.524	2120.4E6	881.8E6	860.891	630.108 #
Target Compounds							
1) T	Dalapon	2.497	2.557	3273.8E6	1824.2E6	560.318	531.335
2) T	3,5-DICHL...	6.197	6.513	2922.3E6	1185.5E6	698.646	584.150
3) T	4-Nitroph...	6.789	7.057	775.2E6	825.5E6	695.553	565.599m
5) T	DICAMBA	7.176	7.711	9760.1E6	4778.2E6	653.087	596.993
6) T	MCP	7.355	7.819	637.4E6	227.2E6	74.273	57.845
7) T	MCPA	7.498	8.052	728.2E6	295.0E6	89.219	59.316 #
8) T	DICHLORPROP	7.859	8.409	2245.2E6	1133.3E6	740.532	590.139
9) T	2,4-D	8.081	8.725	2003.4E6	1252.2E6	802.490	594.692 #
10) T	Pentachlo...	8.361	9.231	34557.6E6	22086.5E6	711.763	622.381
11) T	2,4,5-TP ...	8.929	9.610	13079.4E6	8929.6E6	792.505	611.866
12) T	2,4,5-T	9.212	10.017	11103.5E6	8289.6E6	836.871	616.401 #
13) T	2,4-DB	9.774	10.576	1484.2E6	903.7E6	818.731	616.890
14) T	DINOSEB	10.943	10.950	8406.8E6	5651.7E6	720.054	551.641
15) T	Picloram	10.761	11.997	10303.9E6	10737.1E6	789.554	515.245m#
16) T	DCPA	11.243	11.978	16163.5E6	13312.8E6	808.071	684.952m

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS040125\
Data File : PS029642.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 01 Apr 2025 15:35
Operator : AR\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Manual Integrations
APPROVED

Reviewed By : Abdul Mirza 04/02/2025
Supervised By : mohammad ahmed 04/03/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 02 03:12:04 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS033125.M
Quant Title : 8080.M
QLast Update : Wed Apr 02 02:48:40 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

