

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS040125\
 Data File : PS029647.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 01 Apr 2025 17:35
 Operator : AR\AJ
 Sample : PB167396BS
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 PB167396BS

Manual Integrations
 APPROVED

Reviewed By : Abdul Mirza 04/02/2025
 Supervised By : mohammad ahmed 04/03/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 02 03:13:25 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS033125.M
 Quant Title : 8080.M
 QLast Update : Wed Apr 02 02:48:40 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	6.995	7.522	1708.8E6	700.3E6	693.798	500.453 #
Target Compounds							
1) T	Dalapon	2.495	2.556	2523.5E6	1432.2E6	431.901	417.156
2) T	3,5-DICHL...	6.194	6.511	2343.5E6	941.8E6	560.251	464.087m
3) T	4-Nitroph...	6.786	7.055	642.4E6	660.2E6	576.406	452.365m
5) T	DICAMBA	7.173	7.709	7228.9E6	3735.1E6	483.713	466.660
6) T	MCPD	7.351	7.817	476.9E6	168.8E6	55.575	42.981
7) T	MCPA	7.494	8.049	558.0E6	227.1E6	68.371	45.664 #
8) T	DICHLORPROP	7.857	8.407	1922.4E6	969.2E6	634.051	504.722
9) T	2,4-D	8.078	8.724	1718.5E6	1080.1E6	688.350	512.931 #
10) T	Pentachlo...	8.358	9.229	29064.2E6	17916.7E6	598.618	504.879
11) T	2,4,5-TP ...	8.926	9.608	10551.3E6	7249.4E6	639.325	496.736
12) T	2,4,5-T	9.209	10.015	8781.6E6	6548.8E6	661.870	486.959 #
13) T	2,4-DB	9.772	10.575	1063.2E6	706.6E6	586.509	482.339
14) T	DINOSEB	10.939	10.947	6888.9E6	4579.8E6	590.049	447.015
15) T	Picloram	10.758	11.995	8234.8E6	9363.2E6	631.004	449.315m#
16) T	DCPA	11.240	11.976	12944.6E6	9250.8E6	647.144	475.960m#

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS040125\
Data File : PS029647.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 01 Apr 2025 17:35
Operator : AR\AJ
Sample : PB167396BS
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
PB167396BS

Manual Integrations
APPROVED

Reviewed By : Abdul Mirza 04/02/2025
Supervised By : mohammad ahmed 04/03/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 02 03:13:25 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS033125.M
Quant Title : 8080.M
QLast Update : Wed Apr 02 02:48:40 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

