

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS040220\
 Data File : PS009551.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 02 Apr 2020 17:26
 Operator : DD\AJ
 Sample : PB127979BL
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 03 03:26:56 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS031120.M
 Quant Title : 8080.M
 QLast Update : Thu Mar 12 07:13:24 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.241	6.930	330.8E6	314.2E6	598.093	572.060
Target Compounds							
1) T	Dalapon	3.000	2.586	15943543	110.9E6	20.445	140.833 #
2) T	3,5-DICHL...	0.000	6.106	0	27514697	N.D.	33.831
7) T	MCPA	7.792	0.000	33897214	0	11.130	N.D. #
8) T	DICHLORPROP	0.000	7.750	0	9498089	N.D.	17.119
10) T	Pentachlo...	8.727	0.000	7540586	0	1.072	N.D. #
11) T	2,4,5-TP ...	0.000	8.859	0	8809628	N.D.	2.955
12) T	2,4,5-T	9.449	9.250	29374166	5343804	11.547	1.810 #
13) T	2,4-DB	0.000	9.642	0	3418153	N.D.	9.004
14) T	DINOSEB	11.187	10.046	22170718	12830139	10.846	5.841 #
15) T	Picloram	0.000	11.018	0	17099060	N.D.	4.199
16) T	DCPA	0.000	11.018	0	17099060	N.D.	4.415



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS040220\
Data File : PS009551.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 02 Apr 2020 17:26
Operator : DD\AJ
Sample : PB127979BL
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 03 03:26:56 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS031120.M
Quant Title : 8080.M
QLast Update : Thu Mar 12 07:13:24 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Dalapon


R.T.: 3.000 min
Delta R.T.: -0.046 min
Response: 15943543
Conc: 20.44 ng/ml

Instrument :
ECD_S
ClientSampleId :




#1 Dalapon

R.T.: 2.586 min
Delta R.T.: -0.040 min
Response: 110886817
Conc: 140.83 ng/ml




#2 3,5-DICHLOROBENZOIC ACID

R.T.: 0.000 min
Exp R.T. : 6.451 min
Response: 0
Conc: N.D.



#2 3,5-DICHLOROBENZOIC ACID


R.T.: 6.106 min
Delta R.T.: 0.027 min
Response: 27514697
Conc: 33.83 ng/ml



#4 2,4-DCAA


R.T.: 7.241 min
Delta R.T.: -0.001 min
Response: 330847131
Conc: 598.09 ng/ml

Instrument :
ECD_S
ClientSampleId :




#4 2,4-DCAA

R.T.: 6.930 min
Delta R.T.: -0.005 min
Response: 314173977
Conc: 572.06 ng/ml




#7 MCPA

R.T.: 7.792 min
Delta R.T.: 0.036 min
Response: 33897214
Conc: 11.13 ug/ml



#7 MCPA


R.T.: 0.000 min
Exp R.T.: 7.420 min
Response: 0
Conc: N.D.



#8 DICHLORPROP


R.T.: 0.000 min
Exp R.T.: 8.129 min
Response: 0
Conc: N.D.

Instrument :
ECD_S
ClientSampleId :




#8 DICHLORPROP

R.T.: 7.750 min
Delta R.T.: 0.017 min
Response: 9498089
Conc: 17.12 ng/ml




#10 Pentachlorophenol

R.T.: 8.727 min
Delta R.T.: 0.055 min
Response: 7540586
Conc: 1.07 ng/ml



#10 Pentachlorophenol


R.T.: 0.000 min
Exp R.T.: 8.461 min
Response: 0
Conc: N.D.



#11 2,4,5-TP (SILVEX)


R.T.: 0.000 min
Exp R.T.: 9.221 min
Response: 0
Conc: N.D.

Instrument :
ECD_S
ClientSampleId :




#11 2,4,5-TP (SILVEX)

R.T.: 8.859 min
Delta R.T.: 0.037 min
Response: 8809628
Conc: 2.96 ng/ml




#12 2,4,5-T

R.T.: 9.449 min
Delta R.T.: -0.056 min
Response: 29374166
Conc: 11.55 ng/ml



#12 2,4,5-T


R.T.: 9.250 min
Delta R.T.: 0.049 min
Response: 5343804
Conc: 1.81 ng/ml



#13 2,4-DB


R.T.: 0.000 min
Exp R.T.: 10.059 min
Response: 0
Conc: N.D.

Instrument :
ECD_S
ClientSampleId :




#13 2,4-DB

R.T.: 9.642 min
Delta R.T.: -0.066 min
Response: 3418153
Conc: 9.00 ng/ml




#14 DINOSEB

R.T.: 11.187 min
Delta R.T.: -0.038 min
Response: 22170718
Conc: 10.85 ng/ml



#14 DINOSEB


R.T.: 10.046 min
Delta R.T.: 0.003 min
Response: 12830139
Conc: 5.84 ng/ml



#15 Picloram


R.T.: 0.000 min
Exp R.T.: 11.037 min
Response: 0
Conc: N.D.

Instrument :
ECD_S
ClientSampleId :




#15 Picloram

R.T.: 11.018 min
Delta R.T.: -0.005 min
Response: 17099060
Conc: 4.20 ng/ml



#16 DCPA

R.T.: 0.000 min
Exp R.T.: 11.489 min
Response: 0
Conc: N.D.



#16 DCPA

R.T.: 11.018 min
Delta R.T.: 0.044 min
Response: 17099060
Conc: 4.41 ng/ml