

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS040325\
 Data File : PS029666.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 03 Apr 2025 09:14
 Operator : AR\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Manual Integrations
 APPROVED

Reviewed By :Abdul Mirza 04/03/2025
 Supervised By :mohammad ahmed 04/05/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 03 11:43:33 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS040225.M
 Quant Title : 8080.M
 QLast Update : Wed Apr 02 23:52:55 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	6.963	7.475	1627.8E6	531.5E6	798.379	779.811
Target Compounds							
1) T	Dalapon	2.463	2.534	2466.2E6	1113.1E6	726.980	701.116
2) T	3,5-DICHL...	6.165	6.470	2236.9E6	701.5E6	743.698	726.490
3) T	4-Nitroph...	6.753	7.010	1026.8E6	499.2E6	730.903	704.063
5) T	DICAMBA	7.140	7.662	6405.0E6	2765.7E6	769.650	748.833
6) T	MCPD	7.318	7.768	433.0E6	134.0E6	79.896	79.377
7) T	MCPA	7.461	7.999	568.7E6	169.8E6	78.236	74.651
8) T	DICHLORPROP	7.821	8.357	1589.3E6	683.5E6	722.596	707.416
9) T	2,4-D	8.042	8.672	1754.2E6	766.2E6	728.011	710.441
10) T	Pentachlo...	8.321	9.174	22578.8E6	13596.3E6	742.602	758.397
11) T	2,4,5-TP ...	8.887	9.552	8857.5E6	5427.3E6	762.570	752.440
12) T	2,4,5-T	9.170	9.957	8905.0E6	5080.3E6	774.049	757.928
13) T	2,4-DB	9.729	10.514	1460.2E6	554.8E6	803.752	747.718
14) T	DINOSEB	10.894	10.887	6416.5E6	3778.6E6	760.061	739.056
15) T	Picloram	10.712	11.928	11712.0E6	8802.1E6	755.590	744.402m
16) T	DCPA	11.195	11.912	10875.4E6	6882.7E6	778.700	777.647m

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS040325\
Data File : PS029666.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 03 Apr 2025 09:14
Operator : AR\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Manual Integrations
APPROVED

Reviewed By : Abdul Mirza 04/03/2025
Supervised By : mohammad ahmed 04/05/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 03 11:43:33 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS040225.M
Quant Title : 8080.M
QLast Update : Wed Apr 02 23:52:55 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

