

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS040423\
 Data File : PS022578.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 04 Apr 2023 09:21
 Operator : AR\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 04 21:59:44 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS032023.M
 Quant Title : 8080.M
 QLast Update : Mon Mar 20 15:46:44 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	6.298	6.719	2104.4E6	1596.5E6	860.905	775.135
Target Compounds							
1) T	Dalapon	2.407	2.480	2125.8E6	2354.4E6	681.648	721.430
2) T	3,5-DICHL...	5.680	5.921	2571.3E6	2342.6E6	690.594	734.606
3) T	4-Nitroph...	6.149	6.350	926.8E6	1004.1E6	676.422	715.179
5) T	DICAMBA	6.433	6.876	7232.7E6	7367.1E6	703.553	746.799
6) T	MCP	6.583	6.974	538.4E6	594.4E6	70.824	76.364
7) T	MCPA	6.705	7.184	724.7E6	772.2E6	69.433	73.499
8) T	DICHLORPROP	7.016	7.496	1816.8E6	1825.9E6	685.527	733.267
9) T	2,4-D	7.222	7.789	2237.7E6	2137.5E6	682.434	722.487
10) T	Pentachlo...	7.436	8.203	25313.8E6	24764.1E6	719.445	754.443
11) T	2,4,5-TP ...	7.964	8.567	10448.3E6	10165.6E6	698.999	746.268
12) T	2,4,5-T	8.225	8.943	11715.6E6	9175.1E6	713.443	729.353
13) T	2,4-DB	8.732	9.447	2142.5E6	1265.1E6	710.544	753.372
14) T	DINOSEB	9.766	9.768	9361.2E6	7108.8E6	722.329	741.182
15) T	Picloram	9.619	10.743	15038.2E6	13766.1E6	605.858	686.007
16) T	DCPA	10.035	10.692	14318.9E6	11802.3E6	694.205	744.051

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS040423\
Data File : PS022578.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Apr 2023 09:21
Operator : AR\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 04 21:59:44 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS032023.M
Quant Title : 8080.M
QLast Update : Mon Mar 20 15:46:44 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

