

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS040423\
 Data File : PS022586.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 04 Apr 2023 16:04
 Operator : AR\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 04 22:02:45 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS032023.M
 Quant Title : 8080.M
 QLast Update : Mon Mar 20 15:46:44 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	6.297	6.719	1821.6E6	1616.2E6	745.225	784.663
Target Compounds							
1) T	Dalapon	2.407	2.481	2154.6E6	2368.5E6	690.892	725.745
2) T	3,5-DICHL...	5.679	5.920	2586.1E6	2355.2E6	694.575	738.549
3) T	4-Nitroph...	6.146	6.350	965.5E6	1029.3E6	704.651	733.107
5) T	DICAMBA	6.432	6.876	7249.6E6	7384.4E6	705.192	748.544
6) T	MCPD	6.584	6.974	537.4E6	595.0E6	70.692	76.441
7) T	MCPA	6.706	7.185	729.7E6	776.4E6	69.914	73.899
8) T	DICHLORPROP	7.016	7.496	1831.0E6	1838.2E6	690.874	738.210
9) T	2,4-D	7.221	7.788	2288.5E6	2182.2E6	697.927	737.592
10) T	Pentachlo...	7.436	8.203	25288.1E6	24683.5E6	718.716	751.989
11) T	2,4,5-TP ...	7.964	8.568	10505.3E6	10194.2E6	702.811	748.372
12) T	2,4,5-T	8.224	8.943	11781.7E6	9400.5E6	717.471	747.266
13) T	2,4-DB	8.731	9.448	2002.2E6	1316.3E6	664.021	783.831
14) T	DINOSEB	9.768	9.769	9361.1E6	7147.4E6	722.318	745.205
15) T	Picloram	9.618	10.743	15887.7E6	14290.9E6	640.084	712.160
16) T	DCPA	10.036	10.693	14409.2E6	11868.8E6	698.584	748.241

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS040423\
Data File : PS022586.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Apr 2023 16:04
Operator : AR\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 04 22:02:45 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS032023.M
Quant Title : 8080.M
QLast Update : Mon Mar 20 15:46:44 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

