

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS040521\
 Data File : PS014467.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 05 Apr 2021 12:09
 Operator : DD\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 05 13:41:04 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS031721.M
 Quant Title : 8080.M
 QLast Update : Thu Mar 18 02:23:19 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.158	6.884	1105.9E6	334.1E6	773.973	708.256
Target Compounds							
1) T	Dalapon	3.003	2.596	986.5E6	354.1E6	633.820	598.043
2) T	3,5-DICHL...	6.383	6.046	1365.1E6	488.1E6	707.101	655.524
3) T	4-Nitroph...	6.992	6.500	550.8E6	189.2E6	737.590	622.985
5) T	DICAMBA	7.347	7.050	4014.6E6	1377.3E6	746.505	687.492
6) T	MCPD	7.513	7.145	294.9E6	112.8E6	70.915	67.765
7) T	MCPA	7.662	7.361	428.3E6	162.3E6	74.451	66.626
8) T	DICHLORPROP	8.033	7.676	958.0E6	341.3E6	723.420	669.856
9) T	2,4-D	8.262	7.973	1095.6E6	380.6E6	743.311	637.479
10) T	Pentachlo...	8.568	8.396	14726.2E6	4634.6E6	739.839	678.601
11) T	2,4,5-TP ...	9.113	8.757	5757.1E6	1881.8E6	747.151	691.661
12) T	2,4,5-T	9.398	9.137	5660.3E6	1808.7E6	727.459	678.396
13) T	2,4-DB	9.946	9.643	1057.1E6	232.5E6	683.617	655.602
14) T	DINOSEB	11.099	9.973	4407.7E6	1300.2E6	771.148	719.517
15) T	Picloram	10.924	10.957	6645.0E6	2189.8E6	636.351	577.218
16) T	DCPA	11.365	10.899	8690.8E6	2868.5E6	737.872	699.305

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS040521\
Data File : PS014467.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 05 Apr 2021 12:09
Operator : DD\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 05 13:41:04 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS031721.M
Quant Title : 8080.M
QLast Update : Thu Mar 18 02:23:19 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

