

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS040524\
 Data File : PS026385.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 06 Apr 2024 00:10
 Operator : AR\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Manual Integrations
 APPROVED

Reviewed By :Abdul Mirza 04/08/2024
 Supervised By :Ankita Jodhani 04/09/2024

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 06 03:26:46 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS040224.M
 Quant Title : 8080.M
 QLast Update : Tue Apr 02 13:37:17 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.162	7.664	2239.8E6	675.0E6	720.697	727.282
Target Compounds							
1) T	Dalapon	2.627	2.665	3177.0E6	1175.4E6	637.883	630.100
2) T	3,5-DICHL...	6.345	6.637	3035.0E6	951.3E6	670.629	674.633
3) T	4-Nitroph...	6.959	7.196	1178.6E6	322.2E6	727.317	731.600
5) T	DICAMBA	7.347	7.859	9103.9E6	2950.6E6	691.917	689.664
6) T	MCPD	7.529	7.965	638.6E6	212.1E6	68.508m	68.821
7) T	MCPA	7.677	8.203	868.4E6	282.9E6	66.931	68.039
8) T	DICHLORPROP	8.049	8.565	2307.1E6	724.5E6	679.118	677.000
9) T	2,4-D	8.276	8.887	2422.5E6	711.2E6	704.404	715.301
10) T	Pentachlo...	8.572	9.404	34343.4E6	11128.9E6	740.403	703.939
11) T	2,4,5-TP ...	9.146	9.779	13073.4E6	4210.6E6	701.770	705.295
12) T	2,4,5-T	9.436	10.190	13246.1E6	3888.9E6	729.920	732.454
13) T	2,4-DB	10.008	10.752	1780.8E6	440.5E6	738.206	734.706
14) T	DINOSEB	11.205	11.128	9505.5E6	2776.4E6	756.502	766.336
15) T	Picloram	11.014	12.198	17098.0E6	4611.7E6	750.679	766.261
16) T	DCPA	11.502	12.162	16880.1E6	5018.5E6	713.039	713.361

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS040524\
Data File : PS026385.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 Apr 2024 00:10
Operator : AR\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Manual Integrations
APPROVED

Reviewed By : Abdul Mirza 04/08/2024
Supervised By : Ankita Jodhani 04/09/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 06 03:26:46 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS040224.M
Quant Title : 8080.M
QLast Update : Tue Apr 02 13:37:17 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

