

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS040620\
 Data File : PS009581.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 06 Apr 2020 13:31
 Operator : DD\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 06 17:46:45 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS031120.M
 Quant Title : 8080.M
 QLast Update : Thu Mar 12 07:13:24 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.242	6.930	462.4E6	379.2E6	835.834	690.508
Target Compounds							
1) T	Dalapon	3.046	2.626	519.8E6	500.7E6	666.614	635.966
2) T	3,5-DICHL...	6.451	6.076	581.2E6	524.3E6	769.956	644.678
3) T	4-Nitroph...	7.071	6.532	192.1E6	208.0E6	804.819	653.656
5) T	DICAMBA	7.435	7.101	1633.1E6	1418.5E6	767.387	617.935
6) T	MCPD	7.605	7.198	129.4E6	119.0E6	52.573	65.448
7) T	MCPA	7.755	7.415	183.0E6	170.9E6	60.073	63.317
8) T	DICHLORPROP	8.128	7.728	396.7E6	359.9E6	769.268	648.681
9) T	2,4-D	8.357	8.021	415.3E6	431.0E6	787.624	655.650
10) T	Pentachlo...	8.671	8.455	5349.8E6	4163.6E6	760.862	548.569 #
11) T	2,4,5-TP ...	9.220	8.815	2153.4E6	1875.7E6	785.585	629.178
12) T	2,4,5-T	9.505	9.194	2023.6E6	1876.7E6	795.469	635.692
13) T	2,4-DB	10.057	9.701	281.2E6	257.7E6	812.814	678.966
14) T	DINOSEB	11.222	10.037	1572.2E6	1382.8E6	769.101	629.568
15) T	Picloram	11.036	11.016	2183.8E6	2284.2E6	719.555	560.939
16) T	DCPA	11.487	10.967	2710.3E6	2413.9E6	781.901	623.269

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS040620\
Data File : PS009581.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 Apr 2020 13:31
Operator : DD\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 06 17:46:45 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS031120.M
Quant Title : 8080.M
QLast Update : Thu Mar 12 07:13:24 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

