

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS040621\
 Data File : PS014471.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 06 Apr 2021 08:04
 Operator : DD\AJ
 Sample : HEXANE
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 06 11:28:33 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS031721.M
 Quant Title : 8080.M
 QLast Update : Thu Mar 18 02:23:19 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
----------	------	------	--------	--------	-------	-------

System Monitoring Compounds

Target Compounds							
9) T	2,4-D	8.330f	0.000	3235620	0	2.195	N.D. #
12) T	2,4,5-T	9.441	0.000	8498897	0	1.092	N.D. #
14) T	DINOSEB	11.103	0.000	8217935	0	1.438	N.D. #
15) T	Picloram	10.971	11.001	65428776	24833392	6.266	6.546
16) T	DCPA	11.368	10.904	5325947	2093700	<MDL	<MDL #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS040621\
Data File : PS014471.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 Apr 2021 08:04
Operator : DD\AJ
Sample : HEXANE
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 06 11:28:33 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS031721.M
Quant Title : 8080.M
QLast Update : Thu Mar 18 02:23:19 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#4 2,4-DCAA


R.T.: 0.000 min
Exp R.T. : 7.170 min
Response: 0
Conc: N.D.

Instrument :
ECD_S
ClientSampleId :




#4 2,4-DCAA

R.T.: 0.000 min
Exp R.T. : 6.892 min
Response: 0
Conc: N.D.




#9 2,4-D

R.T.: 8.330 min
Delta R.T.: 0.056 min
Response: 3235620
Conc: 2.20 ng/ml



#9 2,4-D


R.T.: 0.000 min
Exp R.T. : 7.982 min
Response: 0
Conc: N.D.



#12 2,4,5-T


R.T.: 9.441 min
Delta R.T.: 0.030 min
Response: 8498897
Conc: 1.09 ng/ml

Instrument :
ECD_S
ClientSampleId :




#12 2,4,5-T

R.T.: 0.000 min
Exp R.T. : 9.145 min
Response: 0
Conc: N.D.




#14 DINOSEB

R.T.: 11.103 min
Delta R.T.: -0.008 min
Response: 8217935
Conc: 1.44 ng/ml



#14 DINOSEB


R.T.: 0.000 min
Exp R.T. : 9.978 min
Response: 0
Conc: N.D.



#15 Picloram

R.T.: 10.971 min
Delta R.T.: 0.031 min
Response: 65428776
Conc: 6.27 ng/ml

Instrument :
ECD_S
ClientSampleId :



#15 Picloram

R.T.: 11.001 min
Delta R.T.: 0.036 min
Response: 24833392
Conc: 6.55 ng/ml