

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS040623\
 Data File : PS022604.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 06 Apr 2023 16:11
 Operator : AR\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 07 08:21:18 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS032023.M
 Quant Title : 8080.M
 QLast Update : Mon Mar 20 15:46:44 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	6.298	6.719	1861.9E6	1639.8E6	761.698	796.130
Target Compounds							
1) T	Dalapon	2.407	2.481	2115.9E6	2347.4E6	678.465	719.273
2) T	3,5-DICHL...	5.680	5.921	2646.6E6	2416.7E6	710.821	757.830
3) T	4-Nitroph...	6.148	6.351	956.3E6	1036.9E6	697.939	738.510
5) T	DICAMBA	6.432	6.876	7404.3E6	7582.0E6	720.243	768.580
6) T	MCP	6.584	6.974	537.4E6	614.0E6	70.695	78.882
7) T	MCPA	6.706	7.184	735.2E6	796.3E6	70.445	75.796
8) T	DICHLORPROP	7.016	7.496	1862.2E6	1880.9E6	702.641	755.394
9) T	2,4-D	7.222	7.789	2292.8E6	2208.3E6	699.241	746.418
10) T	Pentachlo...	7.436	8.202	25899.2E6	25500.5E6	736.084	776.878
11) T	2,4,5-TP ...	7.964	8.567	10714.6E6	10458.1E6	716.818	767.742
12) T	2,4,5-T	8.225	8.942	12058.8E6	9408.5E6	734.343	747.904
13) T	2,4-DB	8.731	9.447	1880.8E6	1322.0E6	623.746	787.258 #
14) T	DINOSEB	9.767	9.768	9699.6E6	7349.5E6	748.438	766.277
15) T	Picloram	9.619	10.743	15711.1E6	14466.0E6	632.968	720.882
16) T	DCPA	10.035	10.691	14709.3E6	12173.9E6	713.132	767.478

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS040623\
Data File : PS022604.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 Apr 2023 16:11
Operator : AR\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 07 08:21:18 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS032023.M
Quant Title : 8080.M
QLast Update : Mon Mar 20 15:46:44 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

