

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS040623\
 Data File : PS022619.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 07 Apr 2023 00:49
 Operator : AR\AJ
 Sample : 02213-08
 Misc : LOQ WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 LOQ-WATER-02-QT2-2023

Manual Integrations
 APPROVED

Reviewed By : Abdul Mirza 04/07/2023
 Supervised By : Ankita Jodhani 04/07/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 07 08:26:59 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS032023.M
 Quant Title : 8080.M
 QLast Update : Mon Mar 20 15:46:44 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	6.299	6.721	1114.2E6	1175.5E6	455.805	570.719 #
Target Compounds							
1) T	Dalapon	2.408	2.480	403.1E6	680.5E6	129.261m	208.520m#
2) T	3,5-DICHL...	5.683	5.927	646.1E6	592.7E6	173.516	185.867
3) T	4-Nitroph...	6.165	6.365	217.2E6	150.3E6	158.537	107.061 #
5) T	DICAMBA	6.434	6.878	1682.1E6	1691.0E6	163.629	171.415
6) T	MCP	6.577	6.968	126.1E6	135.6E6	16.583m	17.425
7) T	MCPA	6.697	7.176	158.1E6	206.0E6	15.144m	19.612 #
8) T	DICHLORPROP	7.020	7.501	463.5E6	500.8E6	174.904	201.104
9) T	2,4-D	7.239	7.799	500.1E6	586.9E6	152.516	198.373 #
10) T	Pentachlo...	7.437	8.205	7244.7E6	7161.3E6	205.901	218.172
11) T	2,4,5-TP ...	7.966	8.568	2883.6E6	2895.1E6	192.916	212.537
12) T	2,4,5-T	8.230	8.950	3035.6E6	2328.4E6	184.856	185.092
13) T	2,4-DB	8.741	9.456	443.6E6	299.3E6	147.117	178.223
14) T	DINOSEB	9.768	9.770	2775.7E6	1797.3E6	214.180	187.388
15) T	Picloram	9.585	10.759	240.9E6	2972.2E6	9.704m	148.114 #
16) T	DCPA	10.035	10.694	4207.1E6	3087.7E6	203.969	194.654

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS040623\
Data File : PS022619.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 07 Apr 2023 00:49
Operator : AR\AJ
Sample : 02213-08
Misc : LOQ WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
LOQ-WATER-02-QT2-2023

Manual Integrations
APPROVED

Reviewed By : Abdul Mirza 04/07/2023
Supervised By : Ankita Jodhani 04/07/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 07 08:26:59 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS032023.M
Quant Title : 8080.M
QLast Update : Mon Mar 20 15:46:44 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

