

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS040720\
 Data File : PS009630.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 07 Apr 2020 15:52
 Operator : DD\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 07 17:16:59 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS031120.M
 Quant Title : 8080.M
 QLast Update : Thu Mar 12 07:13:24 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.242	6.930	447.8E6	375.3E6	809.565	683.362
Target Compounds							
1) T	Dalapon	3.045	2.627	546.5E6	522.1E6	700.834	663.069
2) T	3,5-DICHL...	6.451	6.076	563.1E6	521.6E6	745.983	641.300
3) T	4-Nitroph...	7.071	6.532	188.4E6	205.4E6	789.240	645.446
5) T	DICAMBA	7.435	7.100	1581.4E6	1418.8E6	743.090	618.063
6) T	MCPD	7.606	7.198	125.3E6	117.5E6	50.935	64.593 #
7) T	MCPA	7.755	7.414	176.5E6	168.3E6	57.955	62.377
8) T	DICHLORPROP	8.129	7.727	382.9E6	356.1E6	742.484	641.782
9) T	2,4-D	8.358	8.021	399.5E6	424.6E6	757.758	645.900
10) T	Pentachlo...	8.671	8.455	5176.7E6	4341.1E6	736.238	571.956
11) T	2,4,5-TP ...	9.220	8.815	2077.3E6	1883.5E6	757.795	631.801
12) T	2,4,5-T	9.504	9.193	1949.6E6	1877.7E6	766.408	636.024
13) T	2,4-DB	10.058	9.701	267.9E6	253.1E6	774.448	666.815
14) T	DINOSEB	11.223	10.036	1476.8E6	1348.9E6	722.464	614.102
15) T	Picloram	11.036	11.015	1881.0E6	2012.5E6	619.778	494.213
16) T	DCPA	11.487	10.966	2588.3E6	2403.1E6	746.705	620.477

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS040720\
Data File : PS009630.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 07 Apr 2020 15:52
Operator : DD\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 07 17:16:59 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS031120.M
Quant Title : 8080.M
QLast Update : Thu Mar 12 07:13:24 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

