

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS040721\
 Data File : PS014512.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 07 Apr 2021 15:06
 Operator : DD\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 07 15:33:54 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS031721.M
 Quant Title : 8080.M
 QLast Update : Thu Mar 18 02:23:19 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.160	6.886	1113.9E6	344.0E6	779.567	729.070
Target Compounds							
1) T	Dalapon	3.005	2.599	1126.1E6	398.5E6	723.497	673.013
2) T	3,5-DICHL...	6.385	6.047	1399.2E6	502.6E6	724.750	674.977
3) T	4-Nitroph...	6.993	6.501	554.2E6	197.2E6	742.054	649.225
5) T	DICAMBA	7.349	7.051	4023.3E6	1407.2E6	748.137	702.452
6) T	MCPD	7.515	7.147	300.7E6	116.9E6	72.322	70.208
7) T	MCPA	7.663	7.362	427.8E6	166.9E6	74.352	68.493
8) T	DICHLORPROP	8.034	7.677	968.3E6	350.1E6	731.243	687.197
9) T	2,4-D	8.263	7.973	1095.6E6	394.2E6	743.298	660.206
10) T	Pentachlo...	8.569	8.397	14806.0E6	4780.6E6	743.847	699.976
11) T	2,4,5-TP ...	9.114	8.758	5744.2E6	1932.6E6	745.478	710.354
12) T	2,4,5-T	9.399	9.138	5747.0E6	1887.7E6	738.606	708.027
13) T	2,4-DB	9.946	9.643	1005.6E6	244.4E6	650.270	689.136
14) T	DINOSEB	11.099	9.973	4417.9E6	1364.1E6	772.948	754.881
15) T	Picloram	10.923	10.956	6894.3E6	2355.8E6	660.226	620.971
16) T	DCPA	11.365	10.900	8643.0E6	2912.7E6	733.811	710.081

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS040721\
Data File : PS014512.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 07 Apr 2021 15:06
Operator : DD\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 07 15:33:54 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS031721.M
Quant Title : 8080.M
QLast Update : Thu Mar 18 02:23:19 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

