



Suggestion Box MSDS RICCA MSDS J.T.Baker
MSDS emdmillipore MSDS restek MSDS CHEMTECH

Welcome Sakhvala
| Logout

Search :

Dashboard | Edd | Efolder | Fax | Genchem | Projecttrack | Qaqc | Quanttrack | Report | Sequence | Utility |

[Home](#) >> [Sequence Module](#) >> [File Info](#)

Sequence Module :: File Info

Navigation Bar

Method : 608.3 PCB

[FileInfo Form](#) [Get Test Info](#) [Analysis RunLog](#) [Sequence Info](#) [Maintenance](#) [Audit Summary Report](#) [Sequence Process](#)

Data For Sequence pq040820 **CHEMTECH-NJ Lab Sequence**
Rule For Ms/Msd of EPH

[Sequence Transfer](#)

[Review Raw Data](#)

[Non Conformance Summary](#)

No Review & No Supervised

Reviewed & No Supervised

Reviewed & Supervised

SampleID	Data File	Date & Time	ClientId	Test	QCType	Dilution.	Comments	Status	System Comments
I.BLK	PQ047311.D	08 Apr 2020 08:10	I.BLK	PCB	IBLK	1	--	OK	UPDATE STATUS
AR1660CCC500	PQ047312.D	08 Apr 2020 08:45	AR1660CCC500	PCB	CCAL500	1	--	OK	UPDATE STATUS
L2151-07	PQ047313.D	08 Apr 2020 09:06	MH-J	PCB	SAM	1	--	OK	UPDATE STATUS
L2151-08	PQ047314.D	08 Apr 2020 09:22	MH-4	PCB	SAM	1	--	OK	UPDATE STATUS
L2151-09	PQ047315.D	08 Apr 2020 09:38	MH-109	PCB	SAM	1	--	OK	UPDATE STATUS
L2151-10	PQ047316.D	08 Apr 2020 09:55	MH-F	PCB	SAM	1	--	OK	UPDATE STATUS
L2151-11	PQ047317.D	08 Apr 2020 10:11	CBH-140	PCB	SAM	1	--	OK	UPDATE STATUS
L2151-12	PQ047318.D	08 Apr 2020 10:27	MH-5	PCB	SAM	1	--	OK	UPDATE STATUS
L2151-13	PQ047319.D	08 Apr 2020 10:44	CBH-594	PCB	SAM	1	--	OK	UPDATE STATUS
L2171-01	PQ047320.D	08 Apr 2020 11:00	FRAC-COMP	PCB	SAM	1	--	OK	UPDATE STATUS
		08							

I.BLK	PQ047321.D	Apr 2020 11:16	I.BLK	PCB	IBLK	1	--	OK	UPDATE STATUS
AR1660CCC500	PQ047322.D	08 Apr 2020 11:32	AR1660CCC500	PCB	CCAL500	1	--	OK	UPDATE STATUS
L2079-15	PQ047323.D	08 Apr 2020 13:29	S6-05-C	PCB Group1	SAM	1	AR1254 & AR1268 Hits, Need 2x Dilution, End CCC fail low	NotOK	Surr, OR Fail UPDATE STATUS
L2123-01	PQ047324.D	08 Apr 2020 13:45	TR-05-040320	PCB	SAM	1	End CCC fail low	NotOK	UPDATE STATUS
L2123-01MS	PQ047325.D	08 Apr 2020 14:01	TR-05-040320MS	PCB	MS	1	End CCC fail low	NotOK	UPDATE STATUS
L2123-01MSD	PQ047326.D	08 Apr 2020 14:54	TR-05-040320MSD	PCB	MSD	1	End CCC fail low	NotOK	UPDATE STATUS
L2172-01	PQ047327.D	08 Apr 2020 15:11	CC-1	PCB	SAM	1	Looks like hexane, End CCC fail low	NotOK	Surr Fail UPDATE STATUS
L2079-08	PQ047328.D	08 Apr 2020 15:27	S6-03-B	PCB Group1	SAM	1	AR1254 & AR1268 Hits, Need 5X,DCB high in both column, End CCC fail low	NotOK	Surr, OR Fail UPDATE STATUS
L2079-09	PQ047329.D	08 Apr 2020 15:43	S6-04-B	PCB Group1	SAM	1	AR1254 & AR1268 Hits, Need 5X,DCB high in both column, End CCC fail low	NotOK	Surr, OR Fail UPDATE STATUS
L2079-10	PQ047330.D	08 Apr 2020 16:00	S6-05-B	PCB Group1	SAM	1	AR1254 & AR1268 Hits, Need 2X,DCB high in both column, End CCC fail low	NotOK	Surr, OR Fail UPDATE STATUS
L2079-13	PQ047331.D	08 Apr 2020 16:16	S6-03-C	PCB Group1	SAM	1	AR1254 & AR1268 Hits,DCB high in both column, End CCC fail low	NotOK	Surr Fail UPDATE STATUS
L2079-15DL	PQ047332.D	08 Apr 2020 16:32	S6-05-CDL	PCB Group1	SAM	2	AR1254 & AR1268 Hits,DCB high in both column, End CCC fail low	NotOK	Surr Fail UPDATE STATUS
L2081-11	PQ047333.D	08 Apr 2020 16:49	S8-01-B	PCB Group1	SAM	1	DCB low in both column, End CCC fail low	NotOK	Surr Fail UPDATE STATUS
L2081-12	PQ047334.D	08 Apr 2020 17:05	S8-02-B	PCB Group1	SAM	1	DCB low in both column, End CCC fail low	NotOK	Surr Fail UPDATE STATUS
L2081-13	PQ047335.D	08 Apr 2020 17:21	S8-03-B	PCB Group1	SAM	1	End CCC fail low	NotOK	UPDATE STATUS

L2081-14	PQ047336.D	08 Apr 2020 17:38	S8-04-B	PCB Group1	SAM	1	AR1268 Hits, End CCC fail low	NotOK	UPDATE STATUS
L2081-15	PQ047337.D	08 Apr 2020 17:54	S8-05-B	PCB Group1	SAM	1	AR1254 & AR1268 Hits, End CCC fail low	NotOK	UPDATE STATUS
L2081-16	PQ047338.D	08 Apr 2020 18:10	S8-06-B	PCB Group1	SAM	1	AR1254 & AR1268 Hits, Need 2X,DCB high in both column, End CCC fail low	NotOK	Surr, OR Fail UPDATE STATUS
L2081-17	PQ047339.D	08 Apr 2020 18:27	S8-07-B	PCB Group1	SAM	1	AR1254 & AR1268 Hits,DCB high in both column, End CCC fail low	NotOK	Surr Fail UPDATE STATUS
L2081-18	PQ047340.D	08 Apr 2020 18:43	S8-08-B	PCB Group1	SAM	1	AR1254 & AR1268 Hits, Need 20X,DCB high in both column, End CCC fail low	NotOK	Surr, OR Fail UPDATE STATUS
L2081-19	PQ047341.D	08 Apr 2020 18:59	S8-09-B	PCB Group1	SAM	1	AR1254 & AR1268 Hits, End CCC fail low	NotOK	UPDATE STATUS
L2081-20	PQ047342.D	08 Apr 2020 19:16	S8-10-B	PCB Group1	SAM	1	AR1254 hit,DCB high in both column, End CCC fail low	NotOK	Surr Fail UPDATE STATUS
I.BLK	PQ047343.D	08 Apr 2020 19:48	I.BLK	PCB	IBLK	1	--	OK	UPDATE STATUS
AR1660CCC500	PQ047344.D	08 Apr 2020 20:04	AR1660CCC500	PCB	CCAL500	1	CCAL Failing low for DCB & AR1260 in both column	NotOK	CCAL Fail UPDATE STATUS
Read Files		Update		All Manual Integration Report					

[Read QUANT File](#)[EPH Sequence Transfer](#)[Update ClientInfo](#)[Trasfer PH](#)