

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS040821\
 Data File : PS014535.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 08 Apr 2021 16:04
 Operator : DD\AJ
 Sample : I.BLK
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 08 17:01:53 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS040821.M
 Quant Title : 8080.M
 QLast Update : Thu Apr 08 14:34:06 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.160	6.886	668.8E6	249.4E6	439.649	535.749
Target Compounds							
9) T	2,4-D	8.328f	0.000	16272407	0	10.234	N.D. #
14) T	DINOSEB	11.078	0.000	19264332	0	2.990	N.D. #
15) T	Picloram	10.961f	10.988	47526861	3303219	4.248	<MDL #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS040821\
Data File : PS014535.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 08 Apr 2021 16:04
Operator : DD\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 08 17:01:53 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS040821.M
Quant Title : 8080.M
QLast Update : Thu Apr 08 14:34:06 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#4 2,4-DCAA


R.T.: 7.160 min
Delta R.T.: 0.002 min
Response: 668791983
Conc: 439.65 ng/ml

Instrument :
ECD_S
ClientSampleId :




#4 2,4-DCAA

R.T.: 6.886 min
Delta R.T.: 0.002 min
Response: 249367097
Conc: 535.75 ng/ml




#9 2,4-D

R.T.: 8.328 min
Delta R.T.: 0.067 min
Response: 16272407
Conc: 10.23 ng/ml



#9 2,4-D


R.T.: 0.000 min
Exp R.T.: 7.970 min
Response: 0
Conc: N.D.



#14 DINOSEB


R.T.: 11.078 min
Delta R.T.: -0.019 min
Response: 19264332
Conc: 2.99 ng/ml

Instrument :
ECD_S
ClientSampleId :




#14 DINOSEB

R.T.: 0.000 min
Exp R.T. : 9.971 min
Response: 0
Conc: N.D.



#15 Picloram

R.T.: 10.961 min
Delta R.T.: 0.042 min
Response: 47526861
Conc: 4.25 ng/ml



#15 Picloram

R.T.: 10.988 min
Delta R.T.: 0.036 min
Response: 3303219
Conc: N.D.