

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS040821\
 Data File : PS014536.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 08 Apr 2021 16:27
 Operator : DD\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 08 17:02:09 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS040821.M
 Quant Title : 8080.M
 QLast Update : Thu Apr 08 14:34:06 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.159	6.885	1171.7E6	364.0E6	770.253	782.081
Target Compounds							
1) T	Dalapon	3.006	2.601	1266.0E6	453.7E6	702.114	697.052
2) T	3,5-DICHL...	6.385	6.047	1483.2E6	534.2E6	715.815	728.486
3) T	4-Nitroph...	6.993	6.501	568.5E6	203.7E6	699.619	721.398
5) T	DICAMBA	7.347	7.051	4248.2E6	1493.9E6	730.145	738.609
6) T	MCPD	7.514	7.147	323.3E6	123.3E6	74.246	73.204
7) T	MCPA	7.662	7.361	452.2E6	174.3E6	72.661	71.770
8) T	DICHLORPROP	8.033	7.675	1015.8E6	369.5E6	725.884	731.943
9) T	2,4-D	8.262	7.972	1153.4E6	405.2E6	725.398	728.585
10) T	Pentachlo...	8.567	8.395	15350.8E6	4961.0E6	730.407	737.963
11) T	2,4,5-TP ...	9.112	8.757	6025.3E6	2032.8E6	736.011	744.149
12) T	2,4,5-T	9.397	9.137	6067.9E6	1970.4E6	737.638	753.338
13) T	2,4-DB	9.944	9.642	1139.3E6	253.5E6	748.657	730.775
14) T	DINOSEB	11.097	9.971	4593.5E6	1403.3E6	712.871	728.721
15) T	Picloram	10.920	10.954	8420.8E6	2840.7E6	752.639	757.544
16) T	DCPA	11.361	10.897	9122.5E6	3045.8E6	740.338	750.061

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS040821\
Data File : PS014536.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 08 Apr 2021 16:27
Operator : DD\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 08 17:02:09 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS040821.M
Quant Title : 8080.M
QLast Update : Thu Apr 08 14:34:06 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

