

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS040924\
 Data File : PS026392.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 09 Apr 2024 14:43
 Operator : AR\AJ
 Sample : HSTDICC1500
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDICC1500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 09 15:01:42 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS040924.M
 Quant Title : 8080.M
 QLast Update : Tue Apr 09 15:01:33 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.106	7.569	3611.2E6	1253.6E6	1430.886	1461.140
Target Compounds							
1) T	Dalapon	2.591	2.612	5926.5E6	2252.9E6	1282.688	1316.481
2) T	3,5-DICHL...	6.293	6.552	5119.9E6	1793.5E6	1328.841	1373.438
3) T	4-Nitroph...	6.900	7.104	1670.7E6	555.6E6	1384.774	1323.589
5) T	DICAMBA	7.289	7.760	15422.5E6	5537.5E6	1399.281	1425.695
6) T	MCPD	7.475	7.871	1187.3E6	429.6E6	150.339	144.048
7) T	MCPA	7.622	8.107	1617.5E6	554.5E6	142.715	138.395
8) T	DICHLORPROP	7.983	8.459	3673.2E6	1320.4E6	1361.283	1368.019
9) T	2,4-D	8.208	8.777	3697.2E6	1254.3E6	1402.168	1408.146
10) T	Pentachlo...	8.505	9.288	43859.4E6	18280.0E6	1199.112	1326.888
11) T	2,4,5-TP ...	9.072	9.663	19618.8E6	7201.4E6	1374.482	1372.814
12) T	2,4,5-T	9.359	10.073	18897.3E6	6417.8E6	1427.455	1434.570
13) T	2,4-DB	9.928	10.633	2666.9E6	734.0E6	1541.635	1477.272
14) T	DINOSEB	11.115	11.008	13102.3E6	4445.1E6	1414.859	1371.789
15) T	Picloram	10.928	12.070	24462.1E6	6594.6E6	1569.242	1631.466
16) T	DCPA	11.413	12.038	24665.3E6	8109.4E6	1391.076	1368.100

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS040924\
Data File : PS026392.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 09 Apr 2024 14:43
Operator : AR\AJ
Sample : HSTDICC1500
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDICC1500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 09 15:01:42 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS040924.M
Quant Title : 8080.M
QLast Update : Tue Apr 09 15:01:33 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

