

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS040924\
 Data File : PS026409.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 09 Apr 2024 22:00
 Operator : AR\AJ
 Sample : PB160076BSD
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 PB160076BSD

Manual Integrations
 APPROVED

Reviewed By : Abdul Mirza 04/10/2024
 Supervised By : Ankita Jodhani 04/10/2024

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 10 02:27:24 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS040924.M
 Quant Title : 8080.M
 QLast Update : Tue Apr 09 15:03:09 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.102	7.564	1409.5E6	396.3E6	558.476	461.964
Target Compounds							
1) T	Dalapon	2.591	2.614	2279.0E6	837.9E6	493.256	489.596
2) T	3,5-DICHL...	6.291	6.548	1988.1E6	572.1E6	516.003	438.066
3) T	4-Nitroph...	6.897	7.099	637.9E6	203.3E6	528.692	484.260
5) T	DICAMBA	7.285	7.755	5746.1E6	1629.7E6	521.344	419.594
6) T	MCP	7.465	7.859	410.4E6	118.0E6	51.973m	39.547
7) T	MCPA	7.611	8.093	567.4E6	162.8E6	50.059	40.638
8) T	DICHLORPROP	7.979	8.453	1426.8E6	402.4E6	528.785	416.882
9) T	2,4-D	8.206	8.772	1426.0E6	391.1E6	540.816	439.066
10) T	Pentachlo...	8.497	9.282	20392.0E6	6042.0E6	557.515	438.569
11) T	2,4,5-TP ...	9.068	9.658	7713.8E6	2176.5E6	540.425	414.907
12) T	2,4,5-T	9.355	10.067	7306.9E6	1952.8E6	551.948	436.508
13) T	2,4-DB	9.923	10.627	941.4E6	215.4E6	544.221	433.433
14) T	DINOSEB	11.110	11.001	4800.0E6	1414.8E6	518.331m	436.606
15) T	Picloram	10.923	12.064	8335.2E6	1967.2E6	534.706m	486.666
16) T	DCPA	11.409	12.031	9692.9E6	2619.4E6	546.660	441.913

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS040924\
Data File : PS026409.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 09 Apr 2024 22:00
Operator : AR\AJ
Sample : PB160076BSD
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
PB160076BSD

Manual Integrations
APPROVED

Reviewed By : Abdul Mirza 04/10/2024
Supervised By : Ankita Jodhani 04/10/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 10 02:27:24 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS040924.M
Quant Title : 8080.M
QLast Update : Tue Apr 09 15:03:09 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

