

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS041321\
 Data File : PS014614.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 13 Apr 2021 20:22
 Operator : DD\AJ
 Sample : M1966-07MSD
 Misc :
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 MW-123RMSD

Manual Integrations
 APPROVED

Ankita
 4/14/2021 12:01:09 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 14 06:37:56 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS040821.M
 Quant Title : 8080.M
 QLast Update : Thu Apr 08 14:34:06 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.157	6.882	770.2E6	249.9E6	506.334	536.910
Target Compounds							
1) T	Dalapon	3.005	2.607	1118.1E6	410.7E6	620.076m	631.077m
2) T	3,5-DICHL...	6.383	6.044	978.3E6	339.2E6	472.152	462.605
3) T	4-Nitroph...	6.991	6.496	375.3E6	196.2E6	461.852	695.020 #
5) T	DICAMBA	7.346	7.049	2717.6E6	988.1E6	467.079	488.534
6) T	MCPD	7.511	7.144	284.0E6	135.4E6	65.226	80.385m
7) T	MCPA	7.659	7.354	309.0E6	267.4E6	49.663	110.119 #
8) T	DICHLORPROP	8.031	7.673	716.2E6	265.7E6	511.819	526.290
9) T	2,4-D	8.259	7.967	874.5E6	356.8E6	550.025	641.705
10) T	Pentachlo...	8.567	8.394	9894.7E6	3365.9E6	470.799	500.689
11) T	2,4,5-TP ...	9.112	8.755	4211.9E6	1478.7E6	514.504	541.307
12) T	2,4,5-T	9.397	9.135	3871.6E6	1354.1E6	470.647	517.728
13) T	2,4-DB	9.944	9.639	531.1E6	203.0E6	349.024m	585.230 #
14) T	DINOSEB	11.095	9.971	2985.0E6	1080.8E6	463.243m	561.261
15) T	Picloram	10.917	10.952	5479.1E6	1766.7E6	489.718	471.125
16) T	DCPA	11.361	10.897	4803.0E6	1844.0E6	389.785	454.118

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS041321\
Data File : PS014614.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 13 Apr 2021 20:22
Operator : DD\AJ
Sample : M1966-07MSD
Misc :
ALS Vial : 26 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
MW-123RMSD

Manual Integrations
APPROVED

Ankita
4/14/2021 12:01:09 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 14 06:37:56 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS040821.M
Quant Title : 8080.M
QLast Update : Thu Apr 08 14:34:06 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

