

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS041420\
 Data File : PS009725.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 14 Apr 2020 09:39
 Operator : DD\AJ
 Sample : HEXANE
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 14 18:02:29 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS031120.M
 Quant Title : 8080.M
 QLast Update : Thu Mar 12 07:13:24 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
----------	------	------	--------	--------	-------	-------

System Monitoring Compounds

Target Compounds							
2) T	3,5-DICHL...	6.490	0.000	25788174	0	34.166	N.D. #
5) T	DICAMBA	0.000	7.162	0	16465497	N.D.	7.173
6) T	MCPP	0.000	7.162	0	16465497	N.D.	9.053
9) T	2,4-D	8.423	0.000	36232596	0	68.723	N.D. #
13) T	2,4-DB	0.000	9.715	0	4942380	N.D.	13.019
15) T	Picloram	10.998	0.000	7765281	0	2.559	N.D. #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS041420\
Data File : PS009725.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 14 Apr 2020 09:39
Operator : DD\AJ
Sample : HEXANE
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 14 18:02:29 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS031120.M
Quant Title : 8080.M
QLast Update : Thu Mar 12 07:13:24 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#2 3,5-DICHLOROBENZOIC ACID


R.T.: 6.490 min
Delta R.T.: 0.039 min
Response: 25788174
Conc: 34.17 ng/ml

Instrument :
ECD_S
ClientSampleId :




#2 3,5-DICHLOROBENZOIC ACID

R.T.: 0.000 min
Exp R.T. : 6.080 min
Response: 0
Conc: N.D.




#4 2,4-DCAA

R.T.: 0.000 min
Exp R.T. : 7.242 min
Response: 0
Conc: N.D.



#4 2,4-DCAA


R.T.: 0.000 min
Exp R.T. : 6.935 min
Response: 0
Conc: N.D.



#5 DICAMBA


R.T.: 0.000 min
Exp R.T. : 7.435 min
Response: 0
Conc: N.D.

Instrument :
ECD_S
ClientSampleId :




#5 DICAMBA

R.T.: 7.162 min
Delta R.T.: 0.057 min
Response: 16465497
Conc: 7.17 ng/ml




#6 MCP

R.T.: 0.000 min
Exp R.T. : 7.606 min
Response: 0
Conc: N.D.



#6 MCP


R.T.: 7.162 min
Delta R.T.: -0.041 min
Response: 16465497
Conc: 9.05 ug/ml



#9 2,4-D


R.T.: 8.423 min
Delta R.T.: 0.065 min
Response: 36232596
Conc: 68.72 ng/ml

Instrument :
ECD_S
ClientSampleId :




#9 2,4-D

R.T.: 0.000 min
Exp R.T. : 8.027 min
Response: 0
Conc: N.D.




#13 2,4-DB

R.T.: 0.000 min
Exp R.T. : 10.059 min
Response: 0
Conc: N.D.



#13 2,4-DB


R.T.: 9.715 min
Delta R.T.: 0.007 min
Response: 4942380
Conc: 13.02 ng/ml



#15 Picloram

R.T.: 10.998 min
Delta R.T.: -0.039 min
Response: 7765281
Conc: 2.56 ng/ml

Instrument :
ECD_S
ClientSampleId :



#15 Picloram

R.T.: 0.000 min
Exp R.T. : 11.023 min
Response: 0
Conc: N.D.