

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS041420\
 Data File : PS009752.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 14 Apr 2020 20:42
 Operator : DD\AJ
 Sample : L2224-01MS
 Misc :
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 FK-BPM-04-009-ABCDMS

Manual Integrations
 APPROVED

Sohil
 4/15/2020 10:38:04 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 15 05:20:27 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS031120.M
 Quant Title : 8080.M
 QLast Update : Thu Mar 12 07:13:24 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.238	6.927	153.1E6	115.9E6	276.748	211.052m
Target Compounds							
1) T	Dalapon	3.044	2.627	145.4E6	147.5E6	186.396	187.288m
2) T	3,5-DICHL...	6.448	6.075	180.6E6	173.7E6	239.284	213.613
3) T	4-Nitroph...	7.067	6.530	65088899	63930166	272.666	200.913 #
5) T	DICAMBA	7.430	7.097	483.4E6	446.4E6	227.133	194.461m
6) T	MCPD	7.595	7.190	40324915	37488376	16.388m	20.611 #
7) T	MCPA	7.744	7.402	60510191	71294927	19.868	26.419 #
8) T	DICHLORPROP	8.124	7.725	132.2E6	133.0E6	256.332	239.631
9) T	2,4-D	8.353	8.017	168.4E6	173.9E6	319.399	264.476
10) T	Pentachlo...	8.665	8.450	1780.5E6	1665.2E6	253.226	219.395
11) T	2,4,5-TP ...	9.215	8.811	737.6E6	697.5E6	269.065	233.980
12) T	2,4,5-T	9.500	9.190	657.7E6	660.7E6	258.548	223.805
13) T	2,4-DB	10.052	9.697	81530399	92696128	235.668	244.184
14) T	DINOSEB	11.216	10.032	105.8E6	96906579	51.758	44.119
15) T	Picloram	11.029	11.011	705.0E6	700.5E6	232.294	172.025 #
16) T	DCPA	11.482	10.962	798.0E6	873.8E6	230.214	225.610

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS041420\
Data File : PS009752.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 14 Apr 2020 20:42
Operator : DD\AJ
Sample : L2224-01MS
Misc :
ALS Vial : 24 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
FK-BPM-04-009-ABCDMS

Manual Integrations
APPROVED

Sohil
4/15/2020 10:38:04 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 15 05:20:27 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS031120.M
Quant Title : 8080.M
QLast Update : Thu Mar 12 07:13:24 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

